

D.No-66/oms/Wdcl/17  
31-01-2017

R.No. 376/CC  
02/02/17

## Anaesthesia workstation Specification (BPL/Penlon)

Anaesthesia Workstation completes with Anaesthesia gas delivery system, Vaporiser, Circle absorber system with built in Anaesthesia ventilator.

### **Anaesthesia Workstation**

- Should have provision for delivery of oxygen, nitrous oxide and medical air with pressure gauges.
- Should have independent attachments for connecting central gas supply and pin indexed cylinders.
- Should have PIN indexed Yoke system for O<sub>2</sub> & N<sub>2</sub>O, Diameter Indexed Safety System for O<sub>2</sub>, N<sub>2</sub>O.
- Should have Pressure gauges for O<sub>2</sub> & N<sub>2</sub>O supply from Cylinder and for O<sub>2</sub>, N<sub>2</sub>O. The gauges should be colour coded should be conveniently placed for easy viewing.
- Oxygen and Nitrous oxide should be linked to ensure a minimum of 25% oxygen delivery at all times to avoid delivery of hypoxic mixture.
- Should have N<sub>2</sub>O cut off facility if O<sub>2</sub> supply fails.
- Should have Oxygen failure alarm both Visual and Audible. Machine should give high priority alarm when O<sub>2</sub> pressure falls below preset limit. Preferably High priority alarm should be distinguishable from normal alarm by way of colour coded messages or alarm lights (Visual) and separate audible tone (Audible)
- Should have Oxygen Flush facility (25 – 75 LPM). O<sub>2</sub> flush switch should be conveniently placed for easy accessibility.
- Should have back bar (Selectatec compatible) with interlocking facility.
- Should have provision for connecting two vaporiser. Vaporizer must be isolated from the gas flow in the off position and prevent the simultaneous activation of more than one vaporizer. Vaporiser must be temperature, pressure and flow compensated.
- Should have selection switch for Open Circuit and Closed circuit operation.
- Should have leak proof compact and integrated circle absorber with Adjustable Pressure Limiting valve, Airway Pressure Measuring device and Bag/Vent switch for Bag to mechanical ventilation.
- Breathing system should have heater system to avoid water condensation.
- Circle absorber should have provision for attaching Oxygen Sensor. Galvanic Oxygen sensor should be supplied.
- Circle absorber should be easily removable without the help of any tools. Should alert the user in case improper fixing of absorber during start-up.
- Should have single soda lime canister of 1.5 litre capacity. Canister removal and fixing should be easy and takes minimum time.
- Should have provision for bypassing the CO<sub>2</sub> absorbent canister during surgical procedure without leaking the system.
- Should have Anaesthesia Gas Scavenging System port (AGSS)
- Should have ACGO Function (Auxiliary Common Gas Outlet)
- Automated pre-check with help menu at start-up for checking leakage and compliance. Test Bypass facility should be provided for emergency use.

