

# Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT  
J.N. MEDICAL COLLEGE HOSPITAL  
ALIGARH MUSLIM UNIVERSITY,  
ALIGARH

No. Ext 1765/MCH

Dated 05.10.15

M/s.....  
.....  
.....

Sealed Quotations/Bids (One Technical and One Financial separately) are invited from manufacturers or their authorized dealers for the following equipments which should be sealed by the tenderers in separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed, on the terms and conditions printed on Page No. 2.

Quotations should reach this office on or before 26.10.2015 by 03:00 PM.

S. No. Items

01.

### SPECIFICATIONS OF ADVANCED 07 MULTIPARA MONITOR

(Brand: BPL/ Nihon Kohden/ Philips/GE Health Care)

- It should be compact in design with atleast 15" color Touch TFT display with minimum 10 waveforms.
- ECG, NIBP, Respiration, Temperature, SPO2, IBP and ETCO2 measurement for the entire patient size ranging from Neonatal to Adult.
- Should have Large Font Display, OxyCRG and Drug Dose calculations.
- Should have atleast 720 hours of graphical and Tabular trends of all parameters along with 1200 sets of NIBP measurements.
- Monitor should have advanced ST Segment and Arrhythmia analysis of all leads.
- Monitor should be capable to upgrade, Cardiac output, Multigas.
- Should have Li-Ion battery to provide the backup of atleast 4 hours of continuous working.
- Should have electrosurgical interference and Defibrillation isolation.
- Should have connectivity option with Central Nursing Station by wired or wireless connection.
- Should have 2 USB ports and Serial ports for future upgrades.
- Should have FDA (USA) approved, CE Certified.

02.

### SPECIFICATIONS OF ADVANCED 05 MULTIPARA MONITOR

(Brand: BPL/ Nihon Kohden/ Philips/GE Healthcare)

- It should be compact in design with atleast 15" color Touch TFT display with minimum 10 waveforms.
- ECG, NIBP, Respiration, Temperature, SPO2, measurement for the entire patient size ranging from Neonatal to Adult.
- Should have Large Font Display, OxyCRG and Drug Dose calculations.
- Should have atleast 720 hours of graphical and Tabular trends of all parameters along with 1200 sets of NIBP measurements.
- Monitor should have advanced ST Segment and Arrhythmia analysis of all leads.
- Monitor should be capable to upgrade, Cardiac output, Multigas.
- Should have Li-Ion battery to provide the backup of atleast 4 hours of continuous working.
- Should have electrosurgical interference and Defibrillation isolation.
- Should have connectivity option with Central Nursing Station by wired or wireless connection.
- Should have 2 USB ports and Serial ports for future upgrades.
- Should have FDA (USA) approved, CE Certified.

Medical Superintendent  
J.N. Medical College Hospital  
A.M.U. Aligarh  
Medical Superintendent  
J.N. Medical College Hospital  
A.M.U. Aligarh

