

**Department of Oral Pathology/Oral Medicine & Radiology**  
**Dr. Z. A. Dental College, A. M. U., Aligarh.**

**CHAIRMAN**

Ref 1808/OP/OM(DC)20

Dated: 14.01.2020

**ENQUIRY FORM**

M/S \_\_\_\_\_

**Dear Sir,**

Sealed quotations /tenders are invited for the following articles on the terms and conditions printed overleaf. QUOTATION should reach this office on or before 04 **February-2020** by 3:00 P.M. positively.

S.No.	Approximate quantity	Particulars/ Specifications
01	01	<p><b>PHOTODYNAMIC THERAPY UNIT</b></p> <ul style="list-style-type: none"><li><input type="checkbox"/> Unit should comprise of 0.5 Watt diode laser emitting heatless light through 3 dimension or 2 dimension light emitting probes.</li><li><input type="checkbox"/> The wavelength of the laser should be 660 Nm.</li><li><input type="checkbox"/> The 2D probe should come with a metallic distance to maintain the ideal distance between the probe and the intraoral mucosal surface.</li><li><input type="checkbox"/> Probe should be sterile packed and disposable in nature.</li><li><input type="checkbox"/> The product should be used along with sterile photosensitiser (5-aminolevulinic acid(ALA),methylene blue to work in conjunction with the light to oral lesions,.</li><li><input type="checkbox"/> Should be applicable in oral premalignant lesions / conditions &amp; aphthous ulcers.</li><li><input type="checkbox"/> The unit should be battery operated with rechargeable cells.</li></ul> <p>Controller unit for timing of photodynamic therapy:</p> <ul style="list-style-type: none"><li><input type="checkbox"/> The unit should be battery operated with rechargeable cells.</li><li><input type="checkbox"/> With programming capacity of one arch at a time &amp; timer showing 60 seconds of radiation around the lesions.</li><li><input type="checkbox"/> Controller should have membrane key pad.</li><li><input type="checkbox"/> The time controller should indicate 20 sec. timing &amp; through acoustic alarm</li></ul>

*Anshul*

(Dr.Anshul Aggarwal )

Chairperson

Chairman

Deptt. of Oral Pathology/Oral  
Medicine & Radiology  
Dr Z.A. Dental College  
AMU, Aligarh, U P

## Terms and Conditions

1. Goods are required to be dispatched F.O.R, Aligarh Ex-go down/at site by passenger train/goods train/parcel post.
2. Goods will be supplied in the name of the Chairman Department of Oral Pathology /Oral Medicine & Radiology Dr.Z.A.Dental College,Aligarh Muslim University Aligarh.
3. The Department has the right to accept the rates of some or all the articles required.
4. The Department reserves the right to reject any or all the quotations without assigning any reason or to allot full or part of the supply to one more firms.
5. Payment shall be made against the bill.
6. In case goods are not according to the specification the cost of returning them shall be borne by the supplier.
7. The goods have to be supplied within Days
8. The Quotations containing uncalled for remarks are likely to be rejected.
9. The firm registered with the sales tax authority should mention sales tax registration no. and PAN wherever applicable.
10. The discount /rebate admissible, if any may be quoted.
11. The rate of sales tax including surcharge along with concession admissible to educational Institution may be specified
12. Other incidental charges such as packing forwarding, insurance etc.may be indicated clearly.
13. The rates material should be valid for one year from the date of the quotations.
14. The firm must be registered either any Government organization or central purchase office A.M.U.Aligarh for supplying of goods items.



Chairperson  
Chairman

Deptt. of Oral Pathology/Oral  
Medicine & Radiology  
Dr Z.A. Dental College  
AMU, Aligarh, U P