

Notice Calling for Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext 1000/MCH/SHKB

Dated 29-10-2019

M/s.....

Sealed Quotation/Bids (one Technical and one Financial separately) are invited from Manufacturer or their authorized dealers for the following items which should be sealed by the tenderers in separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed, on the terms and conditions as mentioned on page No.02

Quotation should reach this office on or before 12/11/2019 by 12:00Noon. & Tech.Bid will be open on 13/11/2019.

BUBBLE CPAP Machine

Specification:

- A. Should have a LCD based graphic display to show
- B. Should have mounting holder for humidifier.
- C. Should have mounted on mobile stand for ease of moving.
- D. Should be compatibles with flow driven CPAP.
- E. Equipment should also work without oxygen input.
- F. Should work on centralized oxygen and air.
- G. Should have CE certification.
- H. Should be supplied from a manufacturing company having ISO certification.
- I. Should be provided with apnea monitoring function.
- J. Should be able to set alarm automatically by press of a single key.
- K. Oxygen should have accuracy:+ 2%
- L. Oxygen concentration range: 21% to 100% F I O2
- M. The pressure generator near the baby should be autoclaveable.
- N. The unit should be on battery back-up at least 1 hour.
- O. Should be provided with single use nasal prong (Small, Medium, Large)
- P. Should be provided with 10, nos of single use circuit.
- Q. Should be provided with disposable bonnets in size of Small, Medium, Large, 10 nos in each size.

Compressor with Standard Accessories (Optional)

- A. Should be able to switch on by press of a single key.
- B. Should be supply along oxygen hose pipe.
- C. Should have supply along oxygen connector.
- D. Should to open to work on centralized oxygen and air.
- E. Should have mounted on castor wheel to move.
- F. Should have CE certification.
- G. Should be supplied from a Manufacturing company having ISO certification.


Servo Humidifier with Standard Accessories Specification:-

- A. Should be provided with single use chamber.
- B. Should be provided with 10 nos of single use Circuit.
- C. Should be provided with sensor probe.
- D. Should have CE certification.
- E. Should be supplied from a Manufacturing company having ISO certification.
- F. Should have alarming function.

Blender:-


Blender for mixing Air & Oxygen

Continued on page No.02


Medical Superintendent
J.N. Medical College Hospital
A.M.U., ALIGARH

Terms & Conditions:

1. Make, Model, Rate, Warranty & GST detail **MUST BE** clearly mentioned in the quotations.
2. Manufacturers/ Authorised dealer's certificate must be attached with the quotations/Tenders.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered with Directorate General of supplier and Disposals (DGS &D) or in Central purchase office A.M.U Aligarh or MSME.
5. **List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 03 years for supply of offered equipment must be enclosed with the price bid.**
6. Installation and Commissioning will be the responsibility of the supplier & after sale service should be provided.
7. The Equipment should be US FDA/CE European approved product.
8. The Instruments/Equipment might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
9. Comprehensive onsite warranty 05years & detail after sale service backup in warranty period must be clearly mentioned.
10. The Rate of CMC for 2years must be provided separately after warranty.
11. The payment shall be made against bill and satisfactory installation. No advance payment is possible.
12. The hospital reserve the right to reject any or all the tender/Quotations without assigning any reason or to allot full or part of supply to one or more firms.
13. Inclusive GST rates will not be accepted and tender /Quotations will be cancelled.
14. The rate must be quoted in Indian rupees and F.O.R JNMC Hospital.
15. Warranty must be clearly defined to the points, about what will be covered during 05years of warranty.
16. Those companies/firms should apply whose head office in India/service centre especially in NCR Delhi.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., ALIGARH
