

**Notice Calling for Quotation/Tender**

OFFICE OF THE MEDICAL SUPERINTENDENT  
J.N. MEDICAL COLLEGE HOSPITAL  
ALIGARH MUSLIM UNIVERSITY ALIGARH

NIT No.34/2019-20/Urology Units/Ext D.No. *637/mch*

Dated. *6/9/2019*

M/s.....  
.....  
.....

Sealed Tenders/Quotations/Bids (One Technical & One Financial) separately are invited from manufacturers or their authorized dealers for the following **Urology Units/Systems** which should be sealed by the tenderers in separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed, on the terms and conditions mentioned on page No.2

Quotations should reach this office on or before 25.09. 2019 by 12:00 Noon and Technical Bids will be opened on 26.09.2019 at 12:00 Noon

<b>Urology Units/Systems</b>	
1	Intra Corporeal Lithotripter (Lithoclast) with following specifications & Compressor Intra-Corporeal Electro Pneumatic Lithotripsy with compressor <ul style="list-style-type: none"><li>• Effective fragmentation</li><li>• Safety</li><li>• Convenient &amp; Easy handling</li><li>• Delivers single or multiple stokes</li></ul> Includes: <ul style="list-style-type: none"><li>• Hand –piece is provided with the suction to suck the fragmented calculi &amp; keep vision clear and other like probes</li><li>• Probes for URS</li><li>• Probes for PCNL</li><li>• Footwitch</li><li>• Control Unit</li><li>• Cylinder Regulator</li><li>• Hand Piece for probes</li></ul>
2	Uroflowmetry system- electronic with standard accessories <ul style="list-style-type: none"><li>• Fully automatic with auto start and stop. Measurement duration :180 seconds</li><li>• Modes of measurement : auto mode and hesitancy mode</li><li>• Successive test can be taken without the need to empty the jar after every test</li><li>• Nomograms in output report</li><li>• It should calculate following parameters<ul style="list-style-type: none"><li>○ Voiding time</li><li>○ Flow time/delay time</li><li>○ Time to max. flow</li><li>○ Max flow rate</li><li>○ Avg. Flow rate</li><li>○ Voiding volume</li></ul></li><li>• It should have an accuracy of +/-3% throughout the scale</li><li>• The equipment should be built in for automatic printout after the test is complete</li><li>• LCD display for compact uroflowmeter</li><li>• Portable, reliable, economical and compact size</li><li>• Weight base/rotating disc based uroflow device</li><li>• Should comply to CE European/USFDA standard</li></ul> <b>Standard Accessories</b> Weight /rotating disc transducer assembly-01 Urine collection basket 2000ml-01 Micturition Chair-01 Funnel-01 Mains cable-01 Printer-01 Printer Cable-01
3	Electronic/remote operated three sectional Urology Theatre table, C-arm compatible (for cystoscopy, ureteroscopy, TURP, TURBT, PCNL, Nephrectomy) with kidney bridge and other standard accessories and extensions

Contd. On Page No.2

*[Signature]*  
**Medical Superintendent**  
J.N. Medical College Hospital  
A.M.U. Aligarh

