

**NOTICE CALLING FOR REAGENT RENTAL BASIS TENDER**

OFFICE OF THE MEDICAL SUPERINTENDENT  
J.N.MEDICAL COLLEGE HOSPITAL  
Aligarh Muslim University, Aligarh-202002

Date- 17/6/2019

D.No..... 200...../MCH- NIT DRUG STORE

Dear Sirs,

Sealed tenders are invited for the following articles on the terms and conditions overleaf. Tenders are on dual bid system i.e. One technical bid and one financial bid seperately.

**Tenders should reach this office on or before 04.07.2019 by 1.00 P.M**

**AUTOMATED URINE ANALYZER ON REAGENT RENTAL BASIS**

**TO PERFORM AUTOMATED INTEGRATED URINARY CHEMISTRY AND EXAMINATION FOR CELLS AND SEDIMENTS**


**SPECIFICATIONS-**

1. Instrument should be fully automated integrated system for both chemistry and microscopic examination with walk away facility.
2. Instrument should be based on principle of reflective photometry method- image or image process by colour sensor or any other method
3. Instrument should perform complete urinary chemistry profile including creatinine, albumin, protien/creatinine ratio and albumin/creatinine ratio.
4. Instrument should have a throughput of upto 250 samples per hour.
5. Instrument should have facility of continuous oading for true walkaway analysis.
6. Instrument should have a data memory of upto 10000 samples.
7. Instrument should have internal barcode identification.
8. Instrument should have facility for interfacing in LIMS/HIMS should be available and should have access to the external printer also.
9. The control and calibrators should be free of cost and there should not be any hidden cost.

**CELL AND SEDIMENT ANALYSIS**


**SPECIFICATIONS**

1. Instrument should have integrated cell & sediment analyser which should be based on flourescence flow cytometry
2. Instrument should be available with following parameters  
(a)RBC/Nonhemolysed RBC ( b). WBC (c) Epithelial Cells (d) Cast (e) Crystals (f) Bacteria (g) Others
3. Instrument should be compatible for body fluid cell counts.
4. Instrument should provide Scatterogrma & histogram for easy interprtation.
5. Instrument should have facility for flagging.
6. Throughput of urine upto 50 samples/hour and of fluid upto 15 samples/hours
7. Instrument should have data memory upto 10000 samples.
8. Instrument should have internal bar code identification.
9. Instrument should have facility for interfacing in LIMS/HIMS and Instrument should have access to th External printer also.
10. The control and calibrators should be free of cost and there should not be any hidden cost.

  
**Medical Superintendent**  
MEDICAL SUPERINTENDENT  
J N. Medical College Hospital  
A.M.U., ALIGARH

## Terms & Conditions:

1. Make, Model, Rate, Warranty & GST detail **MUST BE** clearly mentioned in the quotations.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN code on each item separately.
3. Manufacturers/ Authorized dealer's certificate must be attached with the quotations/Tenders.
4. Cutting/over writing on quotations/tenders will not be accepted.
5. The firm must be registered with Directorate General of supplier and Disposals (DGS &D) or in Central purchase office A.M.U Aligarh.
6. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
7. Installation and Commissioning will be the responsibility of the supplier & after sale service should be provided.
8. The Equipment should be US FDA/ CE European approved product.
9. The Instruments/Equipment might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. Comprehensive onsite warranty 05years & detail after sale service backup in warranty period must be clearly mentioned.
11. The Rate of CMC for 2years must be provided separately after warranty.
12. The payment shall be made against bill and satisfactory installation. No advance payment is possible.
13. The hospital reserve the right to reject any or all the tender/Quotations without assigning any reason or to allot full or part of supply to one or more firms.
14. Inclusive GST rates will not be accepted and tender /Quotations will be cancelled.
15. The rate must be quoted in Indian rupees and F.O.R JNMC Hospital.



**Medical Superintendent**  
**MEDICAL SUPERINTENDENT**  
**J N. Medical College Hospital**  
**A.M.U., ALIGARH**