

J.N. MEDICAL COLLEGE
AMU, ALIGARH
D.No. 175/MCH
DATE 12/6/2019

**OFFICE OF THE MEDICAL SUPERINTENDENT
J.N MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY ALIGARH**

TO SUBSTITUTE ONE BEARING THE SAME NUMBER & DATE

CORRIGENDUM

FOR

IMMUNOSTAINER

Reference Tender No.JNMCH/Histo lab/24/2019-20
Issued vide D.No. 160/MCH dated: 03/06/2019

Tech. Bids will be opened on 25-06-2019 at 12:00Noon in the Office of the Medical Superintendent, JNMCH, AMU Aligarh

Following corrections be made in the specification on the page as mentioned below:

1. **Page No -3 (Point No.5)**

- Should have throughput of at least 20-30 or more slides at a time.

2. **Page No -3 (Point No.10)**


- Should have a liquid cover slip, and air vortex technology for homogenous binding of the Antibody/or may be have a feature to covering the antibody for preventing evaporation and other than liquid cover slipping.

3. **Page No-3 (Point No.13)**

- Should have facility of Individual programming for each slide with any protocol/or may be have a facility for individual antibody for each slide.

4. **Page No -5 (Comprehensive Warranty)**

- Warranty must be clearly defined to the points, about what will be covered during 05years of warranty.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh