

BB/222/19
10/06/19

BLOOD & COMPONENT BANK
J.N. MEDICAL COLLEGE & HOSPITAL AMU, ALIGARH

Tender under sealed cover are invited from reputed companies are invited for **Multi Channel Tube Sealer** with following specifications. Kindly provide financial & technical bids separately.

Specification –


- A. Automatic sequential sealing.
- B. Variable sealing spacing: 70, 80, 85, 90 mm (or any distance up to 200 mm upon request)
- C. Simple operation, fast sealing, No warm up required.
- D. Automatic tubing detection
- E. LED indicator to show sealing process.
- F. Applicable to a wide range of blood bag tubing.
- G. Head protector / Splash guard for user safety.
- H. Applicable tubing diameter 2-6 mm (0.08-0.24 in)
- I. Sealing time – 0.5-1.2 seconds nominal.
- J. Operating temperature – 0 – 40 °C (32-104 °F).
- K. Storage temperature – 20 – 70 °C (4-158 °F).
- L. Humidity - 30% – 80%.
- M. Power Consumption - Operating: 250 watts.
- Stand by: 10 watts.
- N. Fuse - 3 Amps at 100-250 VAC.

Warranty - 5 Years

Quantity - 1 in No.

Tender should be addressed to Medical Superintendent J.N.M.C.H., AMU Aligarh with kind attention Incharge Blood Bank, J.N.M.C.H. , AMU Aligarh. Tender should reach the office of Medical Superintendent by 22/06/2019.

Thanking you.


(Dr. S.H. Arif)
Prof. & I/C Blood Bank
J.N. Medical College & Hospital
AMU, Aligarh

