

BB/149/19
11/4/19

BLOOD & COMPONENT BANK
J.N. MEDICAL COLLEGE & HOSPITAL AMU, ALIGARH

Tender under sealed cover are invited from reputed companies are invited for **Tube Sealer** with following specifications. Kindly provide financial & technical bids separately.

Specification –


- A. Power Supply** - 100-240 VAC (Universal input).
- B. Input Frequency** - 50/60 Hz.
- C. Operation** - Continuous operation.
- D. Power Consumption** - Operating: 170 W.
Standby: 20 W.
- E. Working Environment** - Temperature: - 5°C to 40 °C.
Relative humidity: - 30% to 75%.
(No Condensation)
Atm Pressure: - 600 to 1060 hPa.
- F. RF Output Power (Effective)** - 20W.
- G. RF Output** - 40.68MHz.
- H. Tube Detection** - Automatic.
- I. Maximum Diameter of Tube that can be sealed** - 6mm.
- J. Sealing time** - 1 – 3 Sec (Adjustable).
- K. Fuse** - 250 V, 3.15A, 5X20mm, Type T.

Warranty - 5 Years

Quantity - 3 in No.

Tender should be addressed to Medical Superintendent J.N.M.C.H., AMU Aligarh with kind attention Incharge Blood Bank, J.N.M.C.H., AMU Aligarh. Tender should reach the office of Medical Superintendent by 25/04/19

Thanking you.


(Dr. S.H. Arif)
Prof. & I/C Blood Bank
J.N. Medical College & Hospital
AMU, Aligarh

