

Notice Calling for Quotation/Tender

DNo. 124/CCCR/19
Dated 05.03.2019

OFFICE OF THE DIRECTOR
CENTRE OF CARDIOLOGY
J.N. MEDICAL COLLEGE, AMU, ALIGARH

M/S _____

Sealed Quotations/Bids are invited from Manufacturer or their authorized dealers for the for the purchase of **HEAD UP TILT TABLE** on the terms and conditions as mentioned. Quotation should reach this office on or before **11/03/2019 by 12:00Noon.**

HEAD UP TILT TABLE (HUT) (01)

Technical Specification.

1. Should be of dimensions not less than 180 cms in length and 60 cms. in breadth.
2. Should have proper cushioning of depth not less than 50mm for patient/subject's comfort.
3. Surface should be lined with material which should be water impairment and non-absorbent.
4. Table should be remote controllable.
5. There should be height adjustability of 25" to 39".
6. Should have Head up tilt facility with a tilting range from 0 to 90 degrees. Tilting should be electrically controllable and should be able to be set at any desired angle as shown by the angle indicator.
7. Safe working load should not be less than 280 lbs.
8. Should operate in standard AC line voltage at 110VAC,60Hz, 3amp.
9. Should be mobile with individual locking casters wheels which can be braked to stabilize the table while performing experiments.
10. Should have an angle indicator for tilt.
11. Should have separate foot supports.
12. Should have patients' safety straps to support patient.
13. Should have standalone multiple parameter monitoring device.
14. Should have computer to store Pts. Screening Data.

OPTIONAL FEATURES:

1. Should have provisions for battery back up to perform tilting, lasting for at least 15 min in case of power interruptions.

Terms & Conditions:

1. Make, Model, Rate, Warranty & GST detail **MUST BE** Clearly mentioned in the quotations.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN code on each item separately.
3. Manufacturers/ Authorized dealer's certificate must be attached with the quotations/Tenders.
4. Cutting/over writing on quotations/tenders will not be accepted.
5. The firm must be registered with Directorate General of supplier and Disposals (DGS &D) or in Central purchase office A.M.U Aligarh.
6. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
7. Installation and Commissioning will be the responsibility of the supplier & after sale service should be provided.
8. The Equipment should be US FDA/ CE European approved product.
9. The Instruments/Equipment might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. Comprehensive onsite warranty 05years & detail after sale service backup in warranty period must be clearly mentioned.
11. The Rate of CMC for 2years must be provided separately after warranty.
12. The payment shall be made against bill and satisfactory installation. No advance payment is possible.
13. The Centre of Cardiology reserve the right to reject any or all the tender/Quotations without assigning any reason or to allot full or part of supply to one or more firms.
14. Inclusive GST rates will not be accepted and tender /Quotations will be cancelled.
15. The rate must be quoted in Indian rupees and F.O.R Centre of Cardiology, JNMC AMU Aligarh.



Director
Centre of Cardiology
Director
Centre of Cardiology
J.N. Medical College
A M U Aligarh