

D.No. 2191/06/19  
26/02/19

## Notice Calling for Quotation/Tender

DEPARTMENT OF OBSTETRICS AND GYNAECOLOGY  
J.N.MEDICAL COLLEGE & HOSPITAL  
ALIGARH MUSLIM UNIVERSITY  
ALIGARH.

Date:26.02.2019

**Sealed Quotation/Bids (one Financial and one Technical should be in two separate sealed envelopes otherwise the quotation will be cancelled) are invited from Manufacturer or their Authorised dealers for the following item the terms and conditions printed overleaf.**

**Quotations should reach the office on or before 07.03.2019 by 04:00 PM.**

### RE-TENDER OF CTG MACHINE

S.No	Specification
1	Should have minimum 10" foldable LCD Screen display
2	Monitor should have inbuilt probe rack for better handling
3	Should be suitable for continuous monitoring of Fetal Heart Rate (Dual), uterine Activity.
4	Should have Minimum 12 Crystal water proof transducer for clear FHR
5	Should have ultrasound transducer frequency 1 MHz for contraction tracking
6	Should have dual channel FHR for twin monitoring.
7	Should have Signal overlap verification alerts when two FHR probe detects Heart rate signal from the same source.
8	Should have facility of Manual and Automatic Fetal movements' detections
9	Should be able to store data continuously for 60 hours
10	Should have inbuilt rechargeable lithium-ion battery with battery back up upto 4 hours in continuous mode
11	Should have inbuilt real time thermal printer and should use Z-fold 150mm thermal paper with speed of 1, 2 & 3 cm per min.
12	Should have audio-visual alarm setting for all parameters
13	Option for review stored data and facility of print out.
14	Should have integrated software to analysis report and can atleast report of following
A	Contractions
B	Base heart Line
C	ACCELERATION>10BPM 10 S
D	ACCELERATION>15BPM 15 S
E	Short term variation
F	Long term variation
15	Should have Facility to connect central nursing station with wireless network
16	Should have RJ 45 port to connect with PC to store or transfer review the data
17	It should be upgradable into IUP and DECG (Upgradation should be clearly written in USFDA certification under indication of use <u>OR</u> needs to be physically demonstrated at the time of demonstration of equipment incase required)
18	Should supplied complete with
A	FHR Probe-2No
B	TOCO Probe-1No
C	Event Marker-1No
D	Belt-3No
E	Thermal paper (150mm)-2 Rolls
F	Manual-1 No
19	Fetal monitor Quoted Model should have US FDA Quality Certification

*Il*  
Chairperson  
Deptt. of Obst. & Gynaecology  
J.N. Medical College  
A.M.U., Aligarh

**Terms and Condition:**

1. Make, Model, Rate, Warranty and GST details MUST BE clearly mentioned in the quotation.
2. Only typed quotation/tenders bearing GSTIN No. on original letter head will be entertained
3. And please provide GST-HSN code on each item separately.
4. Manufacturers/Authorised dealer's certificate must be attached with the quotations/Tenders.
5. Cutting/Over writing on quotations will not be accepted.
6. List of users for the quoted items with contact no. should be provided photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in the last 2 years for supply of offered equipment must be enclosed with price bid.
7. Installation and Commission will be the responsibility of the supplier & after sale services will be provided by the seller.
8. The Equipment should be US FDA/CE European approved product.
9. The Instrument/Equipment may be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. Comprehensive onsite warranty of 05 years & detail after sale service backup in warranty period and after warranty period must be clearly mentioned.
11. The rate of CMC for 02 years after warranty must be provided separately.
12. The payment shall be made against bill and satisfactory installation. No advance payment is possible.
13. The department reserve the right to reject any or all tenders/ Quotations without explaining any reason or to allot full or part of supply to one or more firms.
14. The rate must be mentioned in Indian Rupees.



**Prof. Tamkin Khan**  
**(Chairperson)**

Chairperson  
Dept. of Obst. & Gynaecology  
J.N. Medical College  
A.M.U., Aligarh