

ENQUIRY FORM

**DEPARTMENT OF PETROLEUM STUDIES
FACULTY OF ENGG. & TECH.
Aligarh Muslim University, Aligarh-202002**

To

No. Chairman/... MYA (Humidifier) / IPS / D: 5340 Dated: 11.02.2019

Dear Sir,


Please quote your lowest rates, allowing concessions given to Technical Institutions, if any, for the following article keeping in view the terms and condition overleaf, illustrations / specifications stating delivery time should be given in each case, Quotation should reach this office on or before 25.02.2019 under sealed cover. The envelope should be marked with enquiry number.

Kindly go through it carefully and submit the Quotation through Hardcopy and as well as a Scan signed copy with complete details.

S No	Particular														
1.	Temperature controlled humidity chamber: Temperature & humidity controlled chamber with the provision of membrane testing slot (5x5 inch) at the top of chamber that covered with sealed lid. A part of front door should be transparent for better inside view. Provision of pressure control/ relief device for safety & maintaining humidity at high temperature. Chamber must have temperature controlled magnetic stirrer with probe. <table border="1" style="width: 100%;"><tbody><tr><td>Inner Chamber Size</td><td>12"x12"x12" ft</td></tr><tr><td>Temperature Range</td><td>Room temperature to 90 oC</td></tr><tr><td>Humidity range</td><td>25-100 %</td></tr><tr><td>membrane testing slot size</td><td>6x6 inch</td></tr><tr><td>No of Sealed lid required</td><td>03</td></tr><tr><td>Inner Chamber material</td><td>Inner chamber consists of anodized Aluminium or SS 304</td></tr><tr><td>Outer wall</td><td>M.S. Sheet duly painted.</td></tr></tbody></table>	Inner Chamber Size	12"x12"x12" ft	Temperature Range	Room temperature to 90 oC	Humidity range	25-100 %	membrane testing slot size	6x6 inch	No of Sealed lid required	03	Inner Chamber material	Inner chamber consists of anodized Aluminium or SS 304	Outer wall	M.S. Sheet duly painted.
Inner Chamber Size	12"x12"x12" ft														
Temperature Range	Room temperature to 90 oC														
Humidity range	25-100 %														
membrane testing slot size	6x6 inch														
No of Sealed lid required	03														
Inner Chamber material	Inner chamber consists of anodized Aluminium or SS 304														
Outer wall	M.S. Sheet duly painted.														
2.	40 cm long 0.5 mm diameter wire for testing slot is also required with material of. i. Platinum ii. Silver														

Conditions: Quotation must include the following otherwise it will be rejected.

- i. Rate should be quoted FOR Aligarh.
- ii. Certified Brochure must be enclosed.
- iii. The supply must be made by March 20, 2019 as the Funds will lapse on March 25, 2019.
- iv. If the Order is not complied with then the University will be requested to black list your firm.
- v. The complete customer list (institutions) must be attached.
- vi. The validity of the quotation must be at least for 20 days.
- vii. Please provide the Maximum discount possible to the Academic Institutions.
- viii. 100% payment after installation at the site.
- ix. Goods must be supplied within 15 days, else the order shall be treated as cancelled The period of supply can be extended if so desired/cannot be extended.
- x. Goods should be dispatch under Insurance at Supplier risk
- xi. Warranty One year. Extra Charges to be paid if warranty required is for three year at site
- xii. Certificate in support of registration as supplier with Aligarh Muslim University or DGS & D or any central government organization must be enclosed with the quotation failing which it will not be considered.


CHAIRMAN
Deptt. of Petroleum Studies
Aligarh Muslim University
ALIGARH