

No. 1734 /Path.
Dated 4/2/19

DEPARTMENT OF PATHOLOGY , JNMC, AMU.

TENDER ENQUIRY


4.2.19

Sealed quotations are required (technical and financial bid separately) for small automatic chemistry analyzer for chemical lab Department of Pathology JN Medical College, AMU, Aligarh.

Specification

1. Discrete preferable random access, state sample priority should be available.
2. Throughput – Upto 100 test/ hr without ISE.
Upto 200-300 test/ hr with ISE.
3. Based on absorbance photometry or turbidatory or low selective electrode method.
4. All end point and kinetic principles available.
5. Bidirectional LIS interface should be available.
6. Reagent compartment should be refrigerated.

Last date of submission of quotation is 14.02.2019 by 2:00PM in the office of the Chairperson, Department of Pathology.


Chairperson
CHAIRMAN
DEPARTMENT OF PATHOLOGY
J.N. MEDICAL COLLEGE
A.M.U., ALIGARH