



OFFICE OF THE MEDICAL SUPERINTENDENT
J.N MEDICAL COLLEGE HOSPITAL ALIGARH MUSLIM
UNIVERSITY ALIGARH


D.NO. 1816 Purchase Office

Dated: 16-01-2019

EXTENSION OF LAST DATE OF RECEIPT OF TENDER FOR
ANAESTHESIA WORKSTATION

The last date of submission of Tender No. JNMCH/Ophth/15//2018-19 dated 19-12-2018 Regarding **Anaesthesia Workstation** was 15-01-2019. Due to insufficient bids, the date of said tender has been extended upto 25.01.2019 by 12:00 Noon and Tech. bid will be opened on 28.01.2019 other terms & conditions will remain the same.

Copy of the tender is being attached for ready references.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh

OFFICE OF THE MEDICAL SUPERINTENDENT
J.N MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY ALIGARH

TO SUBSTITUTE ONE BEARING THE SAME NUMBER & DATE

CORRIGENDUM

FOR


ANAESTHESIA WORKSTATION

Reference Tender No. JNMCH/Ophth/15/2018-19
Issued vide D.No. 1599/MCH dated: 19/12/2018

Date of submission of bid extended up to 15/01/2019 &
Tech. Bids will be opened on 16-01-2019 at 1:00PM in the Office of the
Medical Superintendent, JNMCH, AMU Aligarh

Following corrections be made in the specification on the pages as mentioned
below:

1. **Page No -3 (General Requirement)**
 - a) Compact and modular, three gas Anaesthesia workstation with an integrated piston/bellow Ventilator for adult to infant.
2. **Page No -3 (Flow Meter)**
 - Dual Cascade type flow meter tubes for Oxygen & N₂O. Range 20 ml / min to 10 Lit/min. Calibrated in multiple scales. Single tube for air 0ml to 10L/min
3. **Page No-3 (Breathing System)**
 - Should have fresh gas de-coupled semi closed circle absorber system/fresh gas compensation.
 - The system should have integrated breathing system to prevent condensation in breathing system and patient comfort (to prevent delivery of dry fresh gases to lungs or mucocilliary transport of fresh gas)
4. **Page No-3 Anaesthesia Ventilator**
 - Should have minimum screen size of 10".
 - Delete : Should have Desflurane compensation.
5. **Page No.-4 Scope of supply**
 - Trolley with 2/3 drawers
 - Delete: Reusable autoclavable microbial filter (5 nos.)
 - Delete: Disposable patient circuits – (50 pcs)


Medical Superintendent
Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh.

Equipment for Hospital

J.N. MEDICAL COLLEGE HOSPITAL

AMU, ALIGARH

D.No. 1599/MCH

DATED 19/12/2018

Advertised Tender Enquiry

For

Anaesthesia Workstation

at

Jawaharlal Nehru Medical College Hospital AMU
Aligarh.

NIT Issue Date	: 19-12-2018
NIT No.	: JNMCH/Ophth/15/2018-19
Pre-Bid Meeting	29-12-2018 at 2:30PM
Last Date of Submission	: 08-01-2019 at 12:30PM
Bid opening	: 09-01-2019 at 11:AM

Tender documents may be downloaded from university website www.amu.ac.in and www.eprocure.gov.in



Aligarh Muslim University Aligarh

Office of the Medical Superintendent

JN Medical College Hospital

AMU Aligarh 202002

Email: jnmedicalpurchase@gmail.com

Medical Superintendent
J.N. Medical College Hospital
A.M.U. Aligarh