

**OFFICE OF THE MEDICAL SUPERINTENDENT
J.N MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY ALIGARH**

TO SUBSTITUTE ONE BEARING THE SAME NUMBER & DATE

CORRIGENDUM

FOR


ANAESTHESIA WORKSTATION

Reference Tender No. JNMCH/Ophth/15/2018-19
Issued vide D.No. 1599/MCH dated: 19/12/2018

**Date of submission of bid extended up to 15/01/2019 &
Tech. Bids will be opened on 16-01-2019 at 1:00PM in the Office of the
Medical Superintendent, JNMCH, AMU Aligarh**

Following corrections be made in the specification on the pages as mentioned below:

1. **Page No -3 (General Requirement)**
 - a) Compact and modular, three gas Anaesthesia workstation with an integrated piston/bellow Ventilator for adult to infant.
2. **Page No -3 (Flow Meter)**
 - Dual Cascade type flow meter tubes for Oxygen & N₂O. Range 20 ml / min to 10 Lit/min. Calibrated in multiple scales. Single tube for air 0ml to 10L/min
3. **Page No-3 (Breathing System)**
 - Should have fresh gas de-coupled semi closed circle absorber system/fresh gas compensation.
 - The system should have integrated breathing system to prevent condensation in breathing system and patient comfort (to prevent delivery of dry fresh gases to lungs or mucocilliary transport of fresh gas)
4. **Page No-3 Anaesthesia Ventilator**
 - Should have minimum screen size of 10".
 - **Delete :** Should have Desflurane compensation.
5. **Page No.-4 Scope of supply**
 - Trolley with 2/3 drawers
 - Delete:** Reusable autoclavable microbial filter (5 nos.)
 - Delete:** Disposable patient circuits – (50 pcs)


Medical Superintendent
Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh