

D.No :- 1512/Patw
29.12.18

DEPARTMENT OF PATHOLOGY, JNMC, AMU.

Dated: 29.12.2018

Quotation

Quotation in technical and commercial bit is needed for YORCO semi automatic microtome are invited from manufacturer and authorized dealer which should be sealed separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed. On the terms and conditions mentioned below should reached to the office of the Chairperson till 10.01.2019 by 2:PM

Specification modular system with exceptional precision design & manufactured in accordance with international standards.

- Best suited for research and routine work in biological, medical or industrial areas.
- Guaranteed over extension protection and sliding-clutch anti-locking system in the coarse advance.
- Highest precision of sections.

The semiautomatic unit offers:-

- Automatic advance for trimming.
- Motorised advance / return of specimen
- Specimen advance automatic in two speeds.

Technical Data

Ranges of section thickness	:	0.5 μ m to 99 μ m
Increments	:	0.5 μ m steps from 1 μ m to 99 μ m
	:	electronically adjustable
Horizontal specimen advance	:	28mm
Vertical stroke	:	60mm
Specimen size	:	upto 30 mm x 30 mm
Specimen orientation	:	X and Y Axis 80 ⁰
	:	Z Axis 360 ⁰
Specimen advance automatic	:	in 2 speeds
Advance for trimming	:	1-99mm adjustable
Electoral data	:	230 volts, 50 cycles single phase AC.


The equipment is supplied complete with:-

- Casette holder
- Knife holder base
- Standard knife holder
- Object orientation unit
- Section waste tray
- microtome knife 15 cm.


CHAIRMAN
DEPARTMENT OF PATHOLOGY
J.N. MEDICAL COLLEGE
A.M.U., ALIGARH

Terms & Conditions:

1. Make, mode, rate and tax details must be clearly mentioned in the quotations.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN code on each item separately.
3. Manufactures/authorised dealer's certificate must be attached with quotations/tenders.
4. The firm must be registered supplier in central purchase organization (e.g. DGS & D) national small scale industries/central purchase office. AMU, Aligarh.
5. List of users for the quoted item with contact no. should provided, photocopy of purchase order along with terms and conditions of contract received from any govt./public sector institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier and after sales service should provided.
7. Cutting/over writing on quotations/ tenders will not be accepted.
8. The equipment/instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
9. The istruments/equipments should be rust proof, made from surgical grade stainless steel and should be CE European/ US FIDA certification
10. Comprehensive onsite warranty 03 years.


CHAIRMAN
DEPARTMENT OF PATHOLOGY
J.N. MEDICAL COLLEGE
A.M.U., ALIGARH

CHAIRMAN
DEPARTMENT OF PATHOLOGY
J.N. MEDICAL COLLEGE
A.M.U., ALIGARH