

RE-TENDER

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext ...1517.../MCH/SHKB

Dated 06/12/2018

M/s.....


.....
.....
.....
Sealed Quotation/Bids (one Technical and one Financial separately) are invited from Manufacturer or their authorized dealers for the following items which should be sealed by the tenderers in separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed, on the terms and conditions as mentioned on page no.02

Quotation should reach this office on or before 20/12/2018 by 12:00Noon.


Specification of Gross Working Stations

- LCD/LED touch screen.
- Displays formalin concentration temperature, humidity.
- Both visual color change & audible buzzer indication for dangerous levels.
- AC outlet for better comfort of the pathologist.
- Automated blower operation based on formalin concentration.
- Easy data transfer-wireless connectivity between grossing station & any enabled device data can be assessed in word format.
- Digital weighing scale.
- Adjustable height within a 20cm range.
- Made of high grade stainless steel.
- Integrated illumination with UV germicidal light.
- Working area of 5'x2.5'x3' and sides made from Acrylic material.
- Self supporting under carriage with adjustable legs.
- Adjustable set of stools with rest to reduce back strain.
- Dictaphone.
- Inbuilt Camera/ Photography system.
- Magnifying glass/ light magnifier with flexible arms.
- Documents holder for requisition forms, literature etc.
- Magnetic instruments holder.
- Paper towel holder.
- Inch/centimeter ruler.
- Polyethylene cutting board.
- Steel basin with adjustable valve with formalin pour off basin.
- Study perforated insertable work plates.
- Waste bin
- Sprinkler system with adjustable valve for bath sanitation located underneath the perforated working plates.
- Removable organizer bins.
- Safety splash shield
- Drawers
- Size 150x75x90cm.

Continued on page 02


Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh

2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN code on each item separately.
3. Manufacturers/ Authorised dealer's certificate must be attached with the quotations/Tenders.
4. Cutting/over writing on quotations/tenders will not be accepted.
5. The firm must be registered with Directorate General of supplier and Disposals (DGS &D) or in Central purchase office A.M.U Aligarh.
6. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
7. Installation and Commissioning will be the responsibility of the supplier & after sale service should be provided.
8. The Equipment should be US FDA/ CE European approved product.
9. The Instruments/Equipment might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. Comprehensive onsite warranty 05years & detail after sale service backup in warranty period must be clearly mentioned.
11. The Rate of CMC for 2years must be provided separately after warranty.
12. The payment shall be made against bill and satisfactory installation. No advance payment is possible.
13. The hospital reserve the right to reject any or all the tender/Quotations without assigning any reason or to allot full or part of supply to one or more firms.
14. Inclusive GST rates will not be accepted and tender /Quotations will be cancelled.
15. The rate must be quoted in Indian rupees and F.O.R JNMC Hospital.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh