

Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext .../1429.../MCH/SHKB

Dated 20/11/2018

M/s.....

Sealed Quotation/Bids (one Technical and one Financial separately) are invited from Manufacturer or their authorized dealers for the following items which should be sealed by the tenderers in separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed, on the terms and conditions as mentioned below :

Quotation should reach this office on or before 04/12/2018 by 12:00Noon.

S.No.	Specification of Defibrillator
1.	Equipment should be portable type to easy mobility.
2.	Equipment should be Bi-Phasic with AED
3.	Equipment should have rechargeable battery.
4.	Equipment should have inbuilt thermal printer.
5.	Equipment should have energy selection from 2j-300j
6.	Equipment should have detachable type external paddle for Adult & Pediatric both.
7.	Equipment should have facility to measure ECG
8.	Equipment should stores and prints at least 24events.
9.	Equipment should have charging time <10sec
10.	Equipment should record ECG before and after shock
11.	Should have the facility of temporary pacing.

Terms & Conditions:

1. Make, Model, Rate, Warranty & GST detail **MUST BE** clearly mentioned in the quotations.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN code on each item separately.
3. Manufacturers/ Authorised dealer's certificate must be attached with the quotations/Tenders.
4. Cutting/over writing on quotations/tenders will not be accepted.
5. The firm must be registered with Directorate General of supplier and Disposals (DGS &D) or in Central purchase office A.M.U Aligarh.
6. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
7. Installation and Commissioning will be the responsibility of the supplier & after sale service should be provided.
8. The Equipment should be US FDA/ CE European approved product.
9. The Instruments/Equipment might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. Comprehensive onsite warranty 05years & detail after sale service backup in warranty period must be clearly mentioned.
11. The Rate of CMC for 2years must be provided separately after warranty.
12. The payment shall be made against bill and satisfactory installation. No advance payment is possible.
13. The hospital reserve the right to reject any or all the tender/Quotations without assigning any reason or to allot full or part of supply to one or more firms.
14. Inclusive GST rates will not be accepted and tender /Quotations will be cancelled.
15. The rate must be quoted in Indian rupees and F.O.R JNMC Hospital.

Medical Superintendent
Medical Superintendent
J.N. Medical College Hospital
A.M.U Aligarh