

# Re-Advertised Tender Enquiry

R-6827/CC  
13.9.18

OFFICE OF THE SUPERINTENDENT  
J.N. MEDICAL COLLEGE HOSPITAL  
ALIGARH MUSLIM UNIVERSITY,  
ALIGARH

No. Ext 975 /MCH/SHKB

Dated 12/9/2018

M/s.....

Sealed Quotations/Bids are invited from Manufacturer or their authorized dealers for the purchase of Multipara Meter for NICU/SNCU on the terms and conditions as mentioned on page no.2

Quotation should reach this office on or before 26/09/2018 by 12:00Noon.

**Note: Those firms who have submitted their bids earlier in response to NIT No.727/MCH/SHKB dated 11-08-2018 need not apply again.**

## FEATURES FOR MULTIPARA METER FOR NICU/SNCU

- 10.4-inch TFT display
- 4 waveforms
- Touch panel monitor control
- 5 parameters: ECG, RESP, SpO2, NIBP, Temp
- Touch screen provides easy and intuitive operation
- Selectable screen layout
- Enlarged numeric screen
- Powerful review with easy access to historical data
- Full Disclosure Up to 120 hours full disclosure of 1 waveform
- Onscreen operation guide
- Effective ECG monitoring 23 types of arrhythmia detection, ST Measurement and selectable QRS detection (adult, child or neo
- PWTT Triggered NIBP Measurement
- Highly Accurate arrhythmia detection,ec 1 arrhythmia analysis
- Alarm history up to 120 hours
- Visual & audible alarm indicator, Selectable & automatic alarm Monitored measurements.
- 3 to 6 hours of continuous monitoring on battery power
- Including standard accessories
- 3 years warranty/Guarantee

Response for service within 24 hours after complaint.

### **Terms & Conditions:**

1. Make, Model, Rate, Warranty & GST detail **MUST BE** Clearly mentioned in the quotations.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN code on each item separately.
3. Manufacturers/ Authorised dealer's certificate must be attached with the quotations/Tenders.

Continued on page 02

Medical Superintendent  
J.N. Medical College Hospital

A.M.U. Aligarh

4. Cutting/over writing on quotations/tenders will not be accepted.
5. The firm must be registered with Directorate General of supplier and Disposals (DGS &D) or in Central purchase office A.M.U Aligarh.
6. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
7. Installation and Commissioning will be the responsibility of the supplier & after sale service should be provided.
8. The Equipment should be US FDA/ CE European approved product.
9. The Instruments/Equipment might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. Comprehensive onsite warranty 05years & detail after sale service backup in warranty period must be clearly mentioned.
11. The Rate of CMC for 2years must be provided separately after warranty.
12. The payment shall be made against bill and satisfactory installation. No advance payment is possible.
13. The hospital reserve the right to reject any or all the tender/Quotations without assigning any reason or to allot full or part of supply to one or more firms.
14. Inclusive GST rates will not be accepted and tender /Quotations will be cancelled.
15. The rate must be quoted in Indian rupees and F.O.R JNMC Hospital.

  
**Medical Superintendent**  
**J.N. Medical College Hospital**  
**A.M.U., Aligarh**  
