

Notice Calling for Quotation/Tender

R-6478/ce
13.8.18

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext 731...../MCH/SHKB

Dated 11/8/2018

M/s.....


.....
.....
Sealed Quotation/Bids (one Technical and one Financial separately) are invited from Manufacturer or their authorized dealers for the following items which should be sealed by the tenderers in separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed, on the terms and conditions as mentioned on page no.2

Quotation should reach this office on or before 06/09/2018 by 12:00Noon.

Technical Specifications for Syringe Pump

- Should be small, compact and light-weight device having dual display: LCD screen for reading parameters and LED for displaying ratings.
- Weight of the device should not exceed 3 Kgs
- Should work on mains cum battery.
- Should have setting modes for Rate, Time and Drug/Weight
- Should have double CPU functioning to ensure reliability and safety
- Should have anti-reverse function which prevents the upstream flow
- Should work on re-chargeble battery with of 8 Hours at 5 ml/hr
- Should have micro and continuous mode of injection
- Should have at least 3 levels of programmable occlusion pressure
- Should have automatic bolus reduction system to avoid accidental bolus delivery after occlusion incident.
- Should work on syringe sizes: 5ml, 10 ml, 20 ml, 30 ml, 50/60 ml
- Should have a flow rate of 0-1200 ml/hr
- Should have an infusion volume range of 0.1-9999.9 ml
- Should have anti-bolus function
- Should have a KVO Rate of 0.1 - 5 ml/hr
- Should be compatible with all types of syringes
- Should have in-built drug library
- History module: At least 1800 history records and last configuration memory.
- Should have alarms for Near Empty, Occlusion, Low Battery, Power Failure, System error, Clamp error etc.
- The manufacturer should be ISO certified
- Should be upgradable to wireless monitoring
- Should be provided with the following standard accessories:
 - With three (03) year warranty
 - User Manual-1 No
 - Power Cord- 1 No

Continued on Page no.02


Medical Superintendent
J.N. Medical College Hospital
A.M.U. Aligarh

Terms & Conditions:

1. Make, Model, Rate, Warranty & GST detail **MUST BE** Clearly mentioned in the quotations.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN code on each item separately.
3. Manufacturers/ Authorised dealer's certificate must be attached with the quotations/Tenders.
4. Cutting/over writing on quotations/tenders will not be accepted.
5. The firm must be registered with Directorate General of supplier and Disposals (DGS &D) or in Central purchase office A.M.U Aligarh.
6. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
7. Installation and Commissioning will be the responsibility of the supplier & after sale service should be provided.
8. The Equipment should be US FDA/ CE European approved product.
9. The Instruments/Equipment might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. Comprehensive onsite warranty 03years & detail after sale service backup in warranty period must be clearly mentioned.
11. The Rate of CMC for 2years must be provided separately after warranty.
12. The payment shall be made against bill and satisfactory installation. No advance payment is possible.
13. The hospital reserve the right to reject any or all the tender/Quotations without assigning any reason or to allot full or part of supply to one or more firms.
14. Inclusive GST rates will not be accepted and tender /Quotations will be cancelled.
15. The rate must be quoted in Indian rupees and F.O.R JNMC Hospital.


Medical Superintendent
J.N. Medical College Hospital
A.M.U. Aligarh
