

D. No. 173 /Path
Dated 14/5/18

133674
14/5/18

BLOOD & COMPONENT BANK
J.N. MEDICAL COLLEGE & HOSPITAL AMU, ALIGARH

Dated: 14.05.2018


To,

The Director,
Computer Center
AMU, Aligarh

Subject: Uploading of enquiry for asking quotation for **Blood Storage Refrigerator (2-6 °C) 600 Liters** in Blood Bank.

Kindly upload the attached enquiry for asking quotation for **Blood Storage Refrigerator (2-6 °C) 600 Liters** to be used in Blood Bank J.N. Medical College & Hospital AMU Aligarh.

Thanking you.


(Dr. S.H. Arif)
Prof. & I/C Blood Bank
J.N. Medical College
AMU, Aligarh



(Annexure I)
BLOOD & COMPONENT BANK
J.N. MEDICAL COLLEGE & HOSPITAL AMU, ALIGARH


Tender under sealed cover are invited from reputed companies are invited for **Blood Storage Refrigerator (2-6 °C) 600 Liters** with following specifications.

Specification –

- | | | |
|----------------------------|---|---|
| A. Input voltage | - | 230±10%V, 50 Hz |
| B. Voltage Stabilizer | - | Built in 1KVA Stabilizer |
| C. Compressor | - | Hermetically sealed compressor |
| D. Refrigerant | - | R134a (CFC Free) |
| E. Trays | - | Stainless steel slide out trays |
| F. Chamber temperature | - | 4±2°C |
| G. Chart Range | - | 0-40 °C |
| H. Chart Resolution | - | 1°C |
| I. Display Resolution | - | 0.5° |
| J. Method of Recording | - | ink pen on paper chart/pressure pen on pressure sensitive chart |
| K. Battery Backup for TRCU | - | 2hr backup with Rechargeable Sealed Maintenance free leadAcid Battery |
| L. Mounting | - | on castor weels |
| M. Internal Lighting | - | Provided using flicker free CFL |

Tender should reach the office of Chairman Dept. of Pathology, J.N Medical College, AMU Aligarh by 30.05.18

Thanking you.


(Dr. S.H. Arif)
Prof. & I/C Blood Bank
J.N. Medical College
AMU, Aligarh

