

Notice Calling Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext 2388 /MCH/SHAKEB

Dated 14/3/2018

M/s.....

.....

Sealed tender are invited from manufacturers or their authorized dealers for the following Item on the terms and conditions mentioned below:

Quotation should reach this office on or before 24/03/2018 by 3:00 PM.

FULLY AUTOMATIC ARM BPM OPD PURPOSE


Measurement method:	Oscillometric
Pressurization:	Internal micro pump
Pressure Display range:	0 to 299 mmHg
NIBP measurement range	SYS 40-270mmHg, DIA 20-200mmHg,Pulse 30-240mmHg
Measurement Accuracy	Pressure: ± 3 mmHg Pulse: $\pm 5\%$
NIBP Clinical test	BS EN1060-4:2004, BHS:1993, ± 5 mmHg SD:8mmHg
Cuff fastening method	Torque Controlled Belt drive Method(TCBM)
Accuracy, Pulse Rate:	± 5 percent
Clock display:	24 hour WITH DATE & TIME
Display Type:	3-digit display by LED
Cuff:	Automatic Operation
Power:	Switching 100-240V AC 50-60Hz
Dimensions:	241(W) x 324(H) x 390(D)
Weight:	Approx. 5.5 Kgs.
Clock function	Clock display , Date & Time (2015 – 2050)
Printer:	Thermal printer with Auto cutter in monitor
Certificate	USFDA/ European certification/ CE
Measurement	By Both Arms
Data transfer	By Bluetooth/ RS232

Terms & Conditions:

1. Make, Model, Rate, Warranty & GST detail MUST BE Clearly mentioned in the quotations.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN code on each item separately.
3. Manufacturers/ Authorised dealer's certificate must be attached with the quotations/Tenders.
4. Cutting/over writing on quotations/tenders will not be accepted.
5. The firm must be registered supplier in central purchase organization (e.g DGS&D)/National small scale Industries/Central purchase office A.M.U Aligarh.
6. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.

Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh

7. Installation and Commissioning will be the responsibility of the supplier & after sale service should be provided.
8. The Instruments/Equipment should be rust proof, made from surgical grade stainless steel and should be US FDA European CE approved product.
9. The Instruments/Equipment might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. Comprehensive onsite warranty 03years & detail after sale service backup in warranty period must be clearly mentioned.
11. The Rate of CMC for 2years must be provided separately after warranty.
12. The payment shall be made against bill and satisfactory installation. No advance payment is possible.
13. The rate must be quoted in Indian rupees and F.O.R JNMC Hospital.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh
Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh 