

Sealed Quotations/Bids (One technical and one financial separately) are invited from manufactures or their authorized dealers for the following item which should be sealed by the tenders in separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed. On the terms and conditions mentioned on page No. 15.03.2018 by 2:00 PM

Specification of gross station for Histopathology lab in Department of Pathology, J.N. Medical college, AMU, Aligarh.

Product description:

Technical Specification

- Made of high grade stainless steel.
- Continuous air flow along with entire area with back draft fume suction.

Suction.


- Integrated Illumination with UV germicidal light.
- Working area 5x2.55x3 & sides made of acrylic.
- Self supporting under carriage with adjustable legs.
- Adjustable sets of stools with rest to reduce back strain.

Optional

- Dictaphone. (optional)
- Camera holder for railing system.
- Magnifying glass/light magnifier with flexible arms.
- Document holder for requisition forms, literature etc.
- Digital weighing machine
- Magnetic instrument holder (optional)
- Paper towel holder
- Inch/Centimetre ruler
- Polyethylene cutting board.
- Tap with swivel neck
- Study perforated insertable work plates

Terms & Conditions:

1. Make, mode, rate and tax details must be clearly mentioned in the quotations.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN code on each item separately.
3. Manufactures/authorised dealer's certificate must be attached with quotations/tenders.
4. The firm must be registered supplier in central purchase organization (e.g. DGS & D) national small scale industries/central purchase office. AMU, Aligarh.
5. List of users for the quoted item with contact no. should provided, photocopy of purchase order along with terms and conditions of contract received from any govt./public sector institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier and after sales service should provided.
7. Cutting/over writing on quotations/ tenders will not be accepted.
8. The equipment/instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
9. The instruments/equipments should be rust proof, made from surgical grade stainless steel and should be CE European/ US FIDA certification
10. Comprehensive onsite warranty 03 years.


CHAIRMAN
Dept. of Path
J.N.M.C., A.M.U.