

No. Ext 2100 /MCH/SHAKEB

Dated 30/11/2018

M/s.....
.....
.....

Sealed Quotations/Bids (One Technical and One Financial separately) are invited from manufacturers or their authorized dealers for the following Item which should be sealed by the tenderers in separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed, on the terms and conditions mentioned below.

Quotations should reach this office on or before 12.02.2018 by 3:00 P.M.

Mobile/Portable 2.5 KW High Frequency X-Ray System with following Specification:

Specification:

- Multimobil 2.5 High Frequency X-ray Generator with the following output:
- Power Output : 2.5 kW as per IS 7620 (Part-1)
- kV range : 40kV-100kV in 24 Steps
- mA range : 13mA-63mA
- mAs range : 0.32 mAs-200mAs
- Exposure Time : 20 msec to 5secs

Safety Standards: IS: 7620, AERB safety code & IEC 60601

Single tank unit with X-ray Tube:

Stationary anode X-ray tube having 1.5x1.5mm focal Spot.

Collimator:

Manually operated LED collimator

Mobile Stand:


Light weight, low height mobile stand with easy manoeuvrability with Cassette box provided

Power Supply:

Single phase 190v-240v $\pm 10\%$; 50Hz power supply with line resistance < 1.5 ohms
Power Consumption: 3 kVA

Terms & Conditions:

1. Make, Model, Rate and Tax detail MUST BE clearly mentioned in the quotations.
2. Only typed quotations /tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN Code on each item separately.
3. Manufacturers/Authorised dealer's certificate must be attached with Quotations/Tenders.
4. The firm must be registered supplier in Central Purchase Organization (e.g. DGS & D)/National Small Scale Industries/Central Purchase Office A.M.U, Aligarh.
5. List of users for the quoted item with contact no. should provided, Photocopy of purchase order along with terms and conditions of contract received from any Govt./Public Sector Institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier and after sales service should provided.
7. Cutting /Over writing on quotations /tenders will not be accepted.
8. The equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
9. Comprehensive on site warranty 03 Years.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh