

Notice Calling for Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY
ALIGARH

No. Ext ... 1646 /MCH/SHAKEB

Dated. 25/11/2017

M/s.....

.....

.....

Sealed Quotations/Bids (One Technical and One Financial separately) are invited from manufacturers or their authorized dealers for the following Item which should be sealed by the tenderers in separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed, on the terms and conditions mentioned below.

Quotations should reach this office on or before 16.12.2017 by 3:00 P.M.

Specification for Intermittent Pneumatic Compression Device. (Sequential Compression Device)

- Provides Sequential, Gradient and Circumferential Compression around the ankle, calf and then the thigh
- Venous Return of individual patients should be sensed by compression system itself using the technique of air plethysmography.
- Provides Animated Alarm Resolution where animated icons communicate the cause of alarm and remedies for alarms.
- Controller should provide reduced noise by having vibration dampeners and soft overmolding.
- Controller should have USB port make software updates to be easy.
- Controller should have adjustable bedhook which attaches easily and securely to most footboards.
- Choice of three styles like Knee Length, Thigh Length & Foot cuff.
- Battery backup with Heavy duty battery which supplies power for 6-8 hours for uninterrupted Compression.
- Controller should have LCD display to show icons cause of alarm and remedies for alarms.
- Sleeves should be comfortable to address patient discomfort.
- System should accommodate single leg operation or both if needed.
- US FDA Mark Certified Quality Product.
- 25 Pair of Knee Length Medium and 25 pair of thigh length small should need to supply with each pump.
- Consumable price should need to be quoted separately and freeze for next 02 Years.
- Warranty should be of 03 Years and CMC for next 02 Years need to be quoted.

Continued on Page No.2

