

Notice Calling Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext¹⁴⁵⁴...../MCH/SHAKEB

Dated.....^{14/10/2017}.....

M/s.....

.....
.....
.....
Sealed tender are invited from manufacturers or their authorized dealers for the following Item on the terms and conditions mentioned below:

Quotation should reach this office on or before ~~02-11-2017~~ by 3:00 PM.

Technical Specification for Microdebrider Universal Console

- Should have in built user friendly interactive menu and illustrative help guide.
- The various parameters should be able to adjust either form touch screen panel or from the screen panel or from the multifunction foot switch.
- Should be able to connect multiple hand pieces at a time like debrider hand pieces (Up to 5000 RPM in oscillating mode & 30,000 RPM in forward mode),Otologic drills (Up to 80,000 RPM), Stapedectomy Drill (Upto 12000 RPM)
- Console should recognize the various hand pieces and automatically adjust the setting accordingly .
- Should have inbuilt pumps each for irrigation (5Cc/Min to 100Cc/Min) and cooling.
- Should have multifunction ergonomically designed foot control with emission for easy identification.
- Should be able to control speed/Mode,Forward/Reverse toggle,Active hand piece change from the foot control itself.
- Should have option for remote control irrigation to operate form sterile area.
- Should have in built scope lens cleaning system.
- Should have provision to connect facial nerve monitor with the otologic drill.
- Should have the option to work without foot paddle.
- Should have option to connect oscillating, Reciprocating and sagittal saws.

Hand piece :

- Able to work up the speed of 12000RPM in forward rotation and 5000RPM in oscillation mode.
- Have fingertip control to rotate only the tip of the blade up to 360deg.
- Have straight suction path to reduce clogging and allow efficient tissue removal.
- Have integrated blade locking system to lock the blade tip rotation.
- Have integrated side grooves and cable clips to provide better tubing management.
- Have titanium body to avoid rusting.
- Should come with zero,12,40, 60,90 and 120 degree blades.

Terms & Conditions:

- 1.Make, Model, Rate, Warranty & GST detail MUST BE Clearly mentioned in the quotations.
- 2.Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN code on each item separately.
- 3.Manufacturers/ Authorised dealer's certificate must be attached with the quotations/Tenders.
- 4.Cutting/over writing on quotations/tenders will not be accepted.
- 5.The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.


6. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.

7. Installation and Commissioning will be the responsibility of the supplier & after sale service should be provided.

9. The Instruments/Equipment should be rust proof, made from surgical grade stainless steel and should be FDA or European CE, UL or BIS approved product.

10. The Instruments/Equipment might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.

11. Comprehensive onsite warranty 03years & detail after sale service backup in warranty period must be clearly mentioned.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh
Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh

