

# Notice Calling Tender

OFFICE OF THE SUPERINTENDENT  
J.N. MEDICAL COLLEGE HOSPITAL  
ALIGARH MUSLIM UNIVERSITY,  
ALIGARH

No. Ext 1455 /MCH/SHAKEB

Dated 14/10/2017

M/s.....

.....

.....

Sealed tender are invited from manufacturers or their authorized dealers for the following Item on the terms and conditions mentioned below:

Quotation should reach this office on or before 02-11-2017 by 3:00 PM.

## Technical specification for Operating Microscope

1. Tubes, eyepieces-	Straight binocular tube, f=170mm -option : Titable binocular tube, F=170mm or f=200mm -10×or 12.5x wide field eye pieces
Objective lenses	Different focal lengths from f=200mm To f=400mm, graded in steps of 50mm
Magnification	Manual, 5-step magnification changer, Magnification factor g=0.4×-2.5×
Focusing	Manual, using the focus adjusting knob Focusing range : 40mm
Illumination	-Coaxial illumination supply via light guide -2×12V 100W halogen reflector lamp -spare lamp with quick-action lamp charger
Maximum load on suspension arm	2.5 to 7.0kg (complete microscope equipment including accessories and coupling)
Rated voltage	115 VAC (100-120VAC ±10%) 230VAC(220-240VAC±10%)
Power consumption	115VAC max. 2.0A 230 VAC max. 1.0A
Rated frequency	50-60Hz
Fuses fuse link	115V ~T6.3A/H 250V 230V~T3.5A/H 250V

**ADDITIONAL ATTACHMENTS –**  
Flexio still camera adaptor  
Medilive prime video camera  
Medialink 100 Video recorder  
Additional objective focal lens of 400mm

## Terms & Conditions:

1. Make, Model, Rate, Warranty & GST detail MUST BE Clearly mentioned in the quotations.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN code on each item separately.
3. Manufacturers/ Authorised dealer's certificate must be attached with the quotations/Tenders.
4. Cutting/over writing on quotations/tenders will not be accepted.
5. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.

6. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.

7. Installation and Commissioning will be the responsibility of the supplier & after sale service should be provided.

9. The Instruments/Equipment should be rust proof, made from surgical grade stainless steel and should be FDA or European CE, UL or BIS approved product.

10. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.

11. Comprehensive onsite warranty 03years & detail after sale service backup in warranty period must be clearly mentioned.

  
Medical Superintendent  
J.N. Medical College Hospital

A.M.U., Aligarh  
Medical Superintendent  
J.N. Medical College Hospital  
A.M.U., Aligarh

