

# Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT  
J.N. MEDICAL COLLEGE HOSPITAL  
ALIGARH MUSLIM UNIVERSITY,  
ALIGARH

No. Ext 1159 /MCH/SHAKEB

Dated 13/9/2017

M/s.....

.....

.....

Sealed Tender/ Quotations are invited from manufacturers or their authorized dealers for the following Items on the terms and conditions mentioned below:

Quotation should reach this office on or before 05/10/2017 by 03:00 P.M.


## Item

### Suction Machine 1.0 HP

- Housing: Powder Coated Anti Corrosive MS, S.S. Top & on Two Wheels
- Jars: Polycarbonate 2 x 2.5 litres capacity
- Overflow Protection: Mechanical Type
- Tubing -Non Collapsible Suction Tubing
- Vacuum Guage: 2.5 inch, 0-760 mmHg
- Filter: Bacterial Filter Autoclavable / Resuable
- Pump Type: Oil Free Twin Headed Piston Pump
- Capacity: -730mmHg  $\pm$  10 at 55-60 LPM
- Noise Level: <50 dB A  $\pm$  3 Almost whisper
- Motor: Thermally Protected
- Power: 220/230V AC, 50 Hz, 820W
- Valves: Stainless Steel

## Terms & Conditions:

1. Make, Model, Rate and Tax detail **MUST BE** clearly mentioned in the quotation
2. Only typed quotations /tenders bearing GSTIN No. on original letter head will be entertained and please provide GST-HSN Code on each item separately.
3. Manufacturers/Authorised dealer's certificate must be attached with Quotations/Tenders.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should provided, Photocopy of purchase order along with terms and conditions of contract received from any Govt./Public Sector Institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier and after sales service should be provided.
7. Cutting /Over writing on quotations /tenders will not be accepted.
8. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL OR BIS approved product.
9. The equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality
10. Comprehensive on site warranty 03 years.

  
Medical Superintendent  
J.N. Medical College Hospital  
A.M.U., Aligarh