

NOTICE CALLING FOR TENDER

OFFICE OF THE MEDICAL SUPERINTENDENT
J.N.MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY
ALIGARH- 202002

No.Ext. 1120 /MCH

Dated 09/09/2017

M/S

.....

.....

Dear Sirs,

Sealed Tenders are invited for the following articles on the terms and conditions printed over leaf.
Quotations should reach this office on or before **25.09.2017** by **1.00 P.M.**

<u>S.No.</u>	<u>Particulars and specifications</u>
---------------------	--

1

Random Access Fully Automated Biochemistry Analyzer with a throughout of 800 tests per hour and above are required for Trauma Center's Lab 1 with the following Terms & Conditions on the basis of Reagent Rental Agreement

Specifications -

1. Assay Mode : End point, Kinetic, Direct ISE and Immuno turbidimetric.
2. At least 8-10 open channels to perform specialized tests/new tests/drugs.
3. On Board Parameters: Minimum 75 on board parameters at a time.
4. Sample Loading: Continuous Load and Unload, 125 routine samples in Universal sample Tray and 8-10 samples in dedicated STAT LANE.
5. It should use disposable tip for sample aspiration and dispense to eliminate carryover & cross contamination.
6. It should have Clot, bubble, low and high viscosity, thin layer fluid and short sample detection, save the sample clot/bubble management, Liquid level sensing. It should have capacity to check hemolysis, Icteric and Turbidity i.e HIT without hampering throughput and without any consumables.
7. The system should be able to take samples from primary/ secondary tubes, sample cups, pediatric cups
8. No daily Calibration should be required & calibration stability must be lot specific and should be stable for 3-4 months.
9. Reaction Cuvettes: Disposable Cuvettes and/or Dry Slides System.

Terms and conditions:-

1. Instrument preferred with minimum water requirement & no external water plant should be required.
2. Certification: Must have quality certification like USFDA and CE from European Union for equipment as well as for the reagents.
3. Certificate of Traceability for calibrators to be submitted by the vendor.
4. No other consumables will be procured for running the investigations. Any consumables essential for the tests have to be supplied by the vendor free of cost for the entire work load during the validity of the contract.
5. At least two machines are required to meet out daily investigations (approx. 3500 test per day).
6. Mention the minimum & maximum period of contract required for the same.
7. Hospital special price list along with general price list with MRP of reagents is required. Per test rates are also required of all the parameters and there should not be any hidden charges.
8. Other terms & conditions will be decided at the time of agreement by mutual consent.

(MEDICAL SUPERINTENDENT)

MEDICAL SUPERINTENDENT
J.N. Medical College Hospital
A.M.U., ALIGARH

