

# NOTICE CALLING FOR TENDER

OFFICE OF THE SUPERINTENDENT  
J.N. MEDICAL COLLEGE HOSPITAL  
ALIGARH MUSLIM UNIVERSITY,  
ALIGARH

No. Ext 1104 /MCH/SHAKEB

Dated 07/09/2017

M/s.....

Sealed quotations are invited from manufacturers or their authorized dealers for the following Item on the terms and conditions mentioned below:


Quotation should reach this office on or before 18/09/2017 by 3:00 PM.

### Technical specification for LED Single /Double film X-ray view box

- Ultra-thin X ray film illuminator using light.
- It should have a thickness of 29mm.
- It should be suitable for viewing 14''×17'' film.
- The LED light must have a life span of more than 1,00,000 hours.
- It should have easy insertion & removal of the film.
- It should have homogenous illumination more than 95% and maximum intensity of over 10,000 LUX.
- It should have an on-off switch along with digital feathers touch dimmer and a button to set the intensity.
- It should have fully electronic continuous brightness control, with adjustment range of approximately 90% .
- It should have directly connectable to power supply without any external adaptor.
- It should have flicker free high frequency light for reduction of eye strain.
- It should have external fuses for protection against power surge.
- 10 steps digital dimmer facility with step up/step down intensity of 400LUX.
- Automatic film sensor.
- Facility to switch on only the section where the film needs to be viewed.
- Once intensity set, should not go back to factory setting.

### **Terms & Conditions:**

1. Make, Model, Rate, Warranty & GST detail MUST BE Clearly mentioned in the quotations with printed price list of the manufacturing firm.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and Commissioning will be the responsibility of the supplier.
7. After sale service should be provided.
8. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL or BIS approved product.
9. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.

  
Medical Superintendent  
J.N. Medical College Hospital  
A.M.U., Aligarh

Medical Superintendent  
J.N. Medical College Hospital  
A.M.U., Aligarh