

R-2317/CC
31.7.17

Notice Calling for Quotation/Tender

OFFICE OF THE MEDICAL SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY
ALIGARH.

No. Ext 865 /MCH/SHAKEB

Dated 29/7/2017

M/s.....
.....
.....

Sealed Quotations are invited from manufacturers or their authorized dealers for the following item on the terms and conditions mentioned below :-

Quotations should reach this office on or before 19/08/2017 by 03.00PM

Specification of Defibrillator

- Equipment should be portable type to easy mobility.
- Equipment Should be Bi-Phasic with AED
- Equipment should have rechargeable battery.
- Equipment should have inbuilt thermal printer.
- Equipment should have energy selection from 2j-300j
- Equipment should have detachable type external paddle for Adult & pediatric both.
- Equipment should have facility to measure ECG.
- Equipment should stores and prints at least 24 events.
- Equipment should have charging time < 10Sec
- Equipment should record ECG before and after shock.
- Should have the facility of temporary pacing .

Terms and Conditions

1. Make,Model,Rate & Tax detail MUST BE clearly mentioned in the quotation.
2. Only typed quotations/tenders bearing GSTIN . on original letter head will be entertained
3. Cutting /Over writing on quotations /tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh
5. List of users for the quoted item with contact no. should provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public Sector Institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier.
 1. After sale service should be provided.
 2. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL or BIS approved product.
 3. The equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
 4. User/Technical/Maintenance manuals to be supplied in English.
 5. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist.
 6. The job description of the hospital technician and company service engineers should be clearly spelt out.
 7. Comprehensive on site warranty 03 years .
 8. Please mention the GST separately on each items.

Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh.

Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh

Placed