

R-2316/CC
31.7.17

Notice Calling for Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext. 864 /MCH/SHAKEB

Dated. 29/7/2017

M/s.....
.....
.....

Sealed quotations are invited from manufacturers or their authorized dealers for the following Item on the term and conditions as mentioned below:

Quotation should reach this office on or before 19/08/2017 by 03:00 P.M.

Ventouse Machine with the following Specification:

- A controlled high pressure noiseless aspirator having compact design and suitable for all surgical disciplines with convenience, precision and safety
- The unit should be compact for good mobility
- It should be based on maintenance free heat resistant, high tech piston/cylinder technology for long lasting dependable operations
- It should be vibration free and noiseless working in operating theatres as well as in patient work
- It should be suitable for all specialised surgery wherever fast and accurate aspiration is required
- It should have autoclavable double jars (3ltr) of unbreakable PSU material
- It should have easy and rapid change over system from one to another jar
- It should have pump displacement of at least 60ltr/min for more
- It should have option to choose variable flow rate of 40 l, 50 l, 60 l to make surgery more easy with simply touch the clean touch buttons
- It should have vacuum of at least 95- kPa (-713 mmHg) or more
- It should have medical grade silicone cups of 50mm, 60mm
- It should have clean touch technology
- It should have touch buttons with LED indicators without gaps and grooves for dust free
- It should have noise level more than of 37-39 dB at 1 meter distance
- It should have front display vacuum gauge with membrane regulator
- It should also have foot on/off switch integrated into the trolley for hands free operation
- It should be fitted on well balanced mobile trolley with 4 antistatic locking castors.
- It should be USFDA approved
- It should also have certification of ISO 9001, ISO 13485, CE 0123
- Warranty should be of atleast 3 years with 2 years CMC after the expiry of warranty period

Terms & Conditions:

1. Make, Model, Rate & Tax detail MUST BE Clearly mentioned in the quotations with printed price list of the manufacturing firm.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and Commissioning will be the responsibility of the supplier.
7. After sale service should be provided.
8. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL or BIS approved product.
9. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. Comprehensive onsite warranty 03 years+2 years CMC
11. Please mention the GST separately on each item.

Medical Superintendent
J.N. Medical College Hospital
Aligarh Muslim University
Aligarh

SM

Handwritten signature