

R-2269/CC
25/7/2017

Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext 820 /MCH/ (HAKERB)

Dated 24/7/2017

M/s.....

Sealed quotations are invited from manufacturers or their authorized dealers for the supply of Syringe Infusion Pump on the term and conditions as mentioned below on Page No. 02

Quotation should reach this office on or before 14/08/2017 by 03:00 P.M.

SYRINGE INFUSION PUMP

Specifications :

Syringe Section	10ml, 20ml, 30ml, 50/60ml
Volume To be nfused (VTBI)	0-999.9ml/hr.,0.1step
Display Volume Infused	0-999.9ml
Accuracy	±2%
Infusion Rate	0.1-400ml/hr.,0.1 step ; Using 50ml/60ml Syringe 0.1-200ml/hr.,0.1 step ; Using 30ml Syringe 0.1-200ml/hr.,0.1 step ; Using 20ml Syringe 0.1-99.9ml/hr.,0.1 step ; Using 10ml Syringe
Highest Infusion Rate	400ml/hr
Purge Rate	400ml/hr;Using 50ml Syringe 200ml/hr;Using 30ml Syringe 200ml/hr;Using 20ml Syringe 99.9ml/hr;Using 10ml Syringe
Alarms	Syringe Installation Failure Alarm, Occlusion Alarm Infusion -Approaching -empty Alarm, Pause Overtime Alarm, Low-Battery Alarm , Completion of Presentting Volume Alarm, power Interruption Alarm, Motor Failure Alarm, Electrical Circuit Jam Alarm.
Power Consumption :	19VA
Occlusion Alarm Pressure : High	20ml,30ml,50ml/60ml Syringe Spec:90Kpa±30Kpa; 10ml Syringe Spec : 100 Kpa±30Kpa
Occlusion Alarm Pressure : Low	20ml,30ml,50ml/60ml Syringe Spec:90Kpa±30Kpa; 10ml Syringe Spec : 70 Kpa±30Kpa
KVO Rate	0.1ml/h±0.002ml/h
Pause Overtime Alarm :	1min 50s-2min
Power	AC:100V-240V 50Hz/60Hz
Battery Type & Duration	Ni-MH DC12V-14V, about 15 hours at the rate of 5ml/hr.
Working Environment	a) Operating Temperature : ±5°C ~ + 40°C; b) Relative Humadity:20% ~ 90%
Dimensions :	290 x 130 x130mm
Weight :	2.00kg
Standards:	Comply with international EMC standards, EN60601-1, EN60601-2-24,EN60601-1-2. CE Certified

Continued on page No. 02

Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh 50/2

Terms and Conditions

1. Make, Model, Rate and Tax detail MUST BE clearly mentioned in the quotation
2. Only typed quotations /tenders bearing GSTIN No. on original letter will be entertained.
3. Cutting /Over writing on quotations /tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should provided, Photocopy of purchase order along with terms and conditions of contract received from any Govt./Public Sector Institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier.
7. After sale service should be provided.
8. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL OR BIS approved product.
9. The equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality
10. Comprehensive on site warranty 03 years.
11. Please mention the GST separately on each item.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh