

Notice Calling for Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext 783 /MCH/SHAKEB

Dated 20/7/2017

M/s.....

Sealed quotations are invited from manufacturers or their authorized dealers for the following items on the terms and conditions mentioned as below.

Quotation should reach this office on or before 12.08.2017 by 03:00 P.M.

- | S. NO. | ITEMS |
|--------|---|
| 01. | Dressing Trolley (Fully S.S)
Approx Size: 30"x18" x 32"
Frame work made of 25mm x 18G, S.S Tubes
Two S.S shelves made of 1mm thick 304 Grade S.S Sheet
S.S Railing on all four sides of the shelves
Fitted with S.S Bucket & S.S Bowl
Trolley mounted on 100mm dia Castors |
| 02. | Instrument Trolley (Fully S.S)
Approx Size: 30"x18" x 32"
Frame work made of 1" x 18G vertical M.S tubes
Two 22G shelves with four sides railing
The trolley mounted on 100mm dia Castors
S.S Frame with S.S shelves |
| 03. | Bed Side Screen (4 Panels)
Size: 1680H x 2450L mm
CRCA tubular frame made of 19mm x 18G I four section
Legs of 25mm x 18G mounted on 5cm on castors with hooks and spring wire alongwith curtains
Each span 610mm wide
Pre treated and Epoxy Powder Coated |
| 04. | LED View Box (Double View) Indian & Imported |

Terms & Conditions:

1. Make, Model, Rate & Tax detail MUST BE Clearly mentioned in the quotations with printed price list of the manufacturing firm.
2. Only typed quotations/tenders bearing GSTIN No. on original letter head will be entertained.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt./Public sector institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and Commissioning will be the responsibility of the supplier. After sale service must be provided.
7. The authorised dealer certificate must be attached with quotation.
8. The Equipment/Instruments should be rust proof, made from surgical grade stainless steel.
9. The Equipment/Instruments should be FDA or CE, UL or BIS approved product.
10. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
11. After sale service should be provided
12. Comprehensive on site warranty 03 years
13. Please mention the GST separately on each item.

Medical Superintendent
J.N. Medical College Hospital
A.M.U. Aligarh
Medical Superintendent
J.N. Medical College Hospital
A.M.U. Aligarh