

Notice Calling for Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext 782 /MCH/SHAKEB

Dated 20/7/2017

M/s.....

.....

.....

Sealed quotations are invited from manufacturers or their authorized dealers for the following Equipments on the terms and conditions mentioned below.

Quotation should reach this office on or before 12/08/2017 by 03:00 P.M.


Diagnostic -2 Channel Audiometer

Specification:-

1. Air Condition Level range: 10 to 120 dBHL depending on head phone and frequency
Headphone: DD 45
2. Bone conduction Level range: 10 to max 80dBHL depending on frequency BC receiver:
B 71 with headband
3. Free field level range: 10 to max 90 dBHL (Canton CD220)
Test Signals: Pure Tones, pulse and warble tone
4. Test frequencies: 125-8,000Hz, opt. up to 20,000Hz
5. Level Steps: Level steps: 5dB, 2dB or 1dB
6. Masking Signals: narrowband noise, white noise and speech noise
7. Tone tests: HL, UCL, Aided
8. Speech tests: SRT, WRS, UCI
9. Over threshold tests: SISI, Decay, ABLB, MLB, Stenger, Langenbeck
10. Feature: MHA
11. Standards: IEC 601-1, IEC 645-1 class 2, IEC 645-2, type B, According to medical Directive 93/42/EEC
12. PC-interface: USB
13. Display: High resolution color display 640x480 pixels, 5.7 inch/14.5cm
14. Power supply: Mains 100-240V -, 50/60 Hz \pm 10%
15. Optional accessories :
 - TDH 39, HOImco 8130, Insert Phones: optional EAR-3A
 - Loudspeaker Canton CD 220 (Single Unit)
 - Portable Loudspeaker SBC
 - CD Player
 - High Frequency headphones HDA 200
 - Talk back microphone
 - Monitor phone, Monitor headset
 - NOAH audiometry module
 - Carrying case

Terms and Conditions

1. Make, Model, Rate and Tax detail MUST BE clearly mentioned in the quotation
2. Only typed quotations /tenders bearing GSTIN No. on original letter will be entertained.
3. Cutting /Over writing on quotations /tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should provided, Photocopy of purchase order along with terms and conditions of contract received from any Govt./Public Sector Institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier.
7. After sale service should be provided.
8. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL OR BIS approved product.
9. The equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality
10. Comprehensive on site warranty 03 years.
11. Please mention the GST separately on each item.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh