

Notice Calling for Tender

OFFICE OF THE MEDICAL SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY
ALIGARH.

No. Ext 504 /MCH/ZAHEER

Dated 13/6/2017

M/s.....

.....

.....

Scaled Quotations are invited from manufacturers or their authorized dealers for the following Equipment on the terms and conditions mentioned below.

Quotations should reach this office on or before 29-06-2017 by 03:00 PM.


Semi Fowler Bed with 2 Sections Mattress 4" & One Telescopic I.V. Rod.

Specification:

- Over all approx size – 2060mm L x 910 mm W x 600mm H
- Frame work 50mm x 30mm x 16G ERW Rectangular tubes.
- Head section be capable of being raised to the angle of 70° from Horizontal level. The head section slope adjustment will be achieved by separate screw mechanism, the rest of bed surface is fixed.
- Tubes Head & Foot Bows of unequal height with vertical tube support.
- 2 sections top made up of oval perforated 0.9mm CRCA thick M.S. sheet.
- Legs fitted with PVC stamp four location for I.V. rod and fitted with one telescopic I.V. Rod.
- Finish – Pretreated & Epoxy powder coated.
- Mattress suitable for bed made of 10 cm thick 50-D PUF covered with good quality Rexene, closed with Zip at one end.

Terms and Conditions

1. Make, Model, Rate and Tax detail MUST BE clearly mentioned in the quotation
2. Only typed quotations /tenders bearing Sales Tax & Tin No on original letter will be entertained.
3. Cutting /Over writing on quotations /tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should provided, Photocopy of purchase order along with terms and conditions of contract received from any Govt./Public Sector Institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier.
7. After sale service should be provided.
8. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL OR BIS approved product.
9. The equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality
10. Comprehensive on site warranty 03 years


Medical Superintendent
J.N. Medical College Hospi"
A.M.U., Aligarh 