

# Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT  
J.N. MEDICAL COLLEGE HOSPITAL  
ALIGARH MUSLIM UNIVERSITY,  
ALIGARH

No. Ext. 377/MCH /MCH/ZAHEER

Dated. 23/05/2017

M/s.....

.....

.....

Sealed quotations are invited from manufacturers or their authorized dealers for the following items on the terms and conditions mentioned as below.

Quotations should reach this office on or before 06/06/2017 by 05:00 P.M.

- | S. NO. | ITEMS   |
|--------|---|
| 01.    | <b>Dressing Trolley (Fully S.S)</b><br>Approx Size: 30"x18" x 32"<br>Frame work made of 25mm x 18G, S.S Tubes<br>Two S.S shelves made of 1mm thick 304 Grade S.S Sheet<br>S.S Railing on all four sides of the shelves<br>Fitted with S.S Bucket & S.S Bowl<br>Trolley mounted on 100mm dia Castors   |
| 02.    | <b>Instrument Trolley (Fully S.S)</b><br>Approx Size: 30"x18" x 32"<br>Frame work made of 1" x 18G vertical M.S tubes<br>Two 22G shelves with four sides railing<br>The trolley mounted on 100mm dia Castors<br>S.S Frame with S.S shelves  |
| 03.    | <b>Bed Side Screen (4 Panels)</b><br>Size: 1680H x 2450L mm<br>CRCA tubular frame made of 19mm x 18G I four section<br>Legs of 25mm x 18G mounted on 5cm on castors with hooks and spring wire alongwith curtains<br>Each span 610mm wide<br>Pre treated and Epoxy Powder Coated  |
| 04.    | <b>Patient Stretcher on Trolley</b><br>Size: 48" x 22" x 33"<br>Frame work 1 1/4" x 16 G Horizontal & 1 1/2" x 16 G vertical CRCA Tube<br>Removable Stretcher Top made of M.S sheet 18G with S.S handle 1" x 16 G<br>Four noiseless 15cm diameter heavy duty castor wheels, 2Nos with locks for break movement<br>Wheels should not be welded out the trolley as repair/replacement of wheels later on is possible<br>Finish-pre-treated & Epoxy powder coated<br>Provided with on cylinder cage and arrangement for I/V stand<br>Mattress 2" PUF 32-D covered with good quality Rexene cover and laminated |
| 05     | <b>LED view Box (Double View) Indian &amp; Imported</b>   |

#### Terms & Conditions:

1. Rate & Tax detail, Make, Model MUST BE Clearly mentioned in the quotations with printed price list of the manufacturing firm.
2. Only typed quotations/tenders bearing Sales Tax & TIN Nos. on original letter head will be entertained.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt./Public sector institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and Commissioning will be the responsibility of the supplier. After sale service must be provided.
7. The authorised dealer certificate must be attached with quotation.
8. The Equipment/Instruments should be rust proof, made from surgical grade stainless steel.
9. The Equipment/Instruments should be FDA or CE, UL or BIS approved product.
10. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
11. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist and the job description of the hospital technician and company service engineers should be clearly spelt out.
12. Comprehensive on site warranty 03 years

Medical Superintendent  
J.N. Medical College Hospital  
A.M.U., Aligarh

  
