

Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext¹⁹⁵...../MCH/NAZEER

Dated.....^{21/5/2017}.....

M/s.....
.....
.....

Sealed quotations are invited from manufacturers or their authorized dealers for the following Item on the term and conditions as mentioned below:


Quotation should reach this office on or before 13.05.2017 by 03:00 P.M.

- | S. NO. | ITEMS |
|--------|---|
| 1. | Scrub Station (Double User)
Made of heavy duty stainless steel 304 grade
With Foot Cum Auto Sensor Operated
Size:- 145L x 54W x 84H |
| 2. | Crash Cart (Fully S.S)
Overall Approx, Size: 960 x 500W x 1545H mm
Frame work made of S.S. Tubes, Two light weight polystyrene boxes each with three drawers, Upper drawer with medicine container of different sizes, Provision to hold oxygen cylinder made of M.S., Massage board for cardiac patient are provided, Six Number colour hand out small bins at eye level, Total three S.S. Shelves with railing to carry Medical Equipments, Provided with S.S. I.V. Rod, Corner Butters, The cart mounted on 125mm dia castors 2 with brakes |
| 3. | Mobile O.T. Light, (Spot Light)
Single reflector mobile halogen O.T. Light mounted on sturdy stand. Low voltage transformer integrated in the base. Effective heat absorbing, colour correcting filter & shadow reducing diffuser. Each dome consists of two halogen bulbs as standard. Reserve halogen bulb can be switched ON required. Spring balanced, flexible arm positioning without any drift. Intensity Control and Main Switch provided on the front panel. Inbuilt Switch Mode Power Supply (SMPS) / CVT. Smooth white powder coated finish.

Light Intensity 50,000 Lux. Dia. of Dome : 500mm (Approx)
Dichroic Reflector Dimension: 400 mm, Halogen Bulb : 24V 150W |
| 4. | Cardiac Table (Fixed Height)
Size:- L 106 x W 45 x H 95 cm (Approx)
Laminated wooden top size – 30” x 16” |

Terms & Conditions:

1. Rate & Tax detail MUST BE Clearly mentioned in the quotations with printed price list of the manufacturing firm.
2. Only typed quotations/tenders bearing Sales Tax & TIN Nos. on original letter head will be entertained.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt./Public sector institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and Commissioning will be the responsibility of the supplier. After sale service must be provided.
7. The authorised dealer certificate must be attached with quotation.
8. The Equipment/Instruments should be rust proof, made from surgical grade stainless steel.
9. The Equipment/Instruments should be FDA or CE, UL or BIS approved product.
10. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
11. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist and the job description of the hospital technician and company service engineers should be clearly spelt out.
12. Comprehensive on site warranty 03 years.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., ALIGARH