

CORRIGENDUM

Tender No. : JNMC Cath Lab (P) / Tender / 04 / 2017-18 dated 25/3/17
 Pre bid Meeting on 03/04/2017 in the office of Medical Superintendent at 3:00p.m

CHANGES IN SPECIFICATIONS OF CARDIAC SURGERY EQUIPEMENTS

S NO	ITEM	ORIGINAL SPECIFICATIONS	TO BE READ AS
1.	Heart lung machine with Heater Cooler Unit/Temperature control Module(TCM)	No change	
2.	9 parameter monitor with 2 slave monitors	At least 26 inch Color TFT LCD/LED display	At least 19 inch color TFT LCD/LED display
		Touch screen facility	Touch screen and rotatory Knob control facility
		Advanced Masimo Technology available which monitors SpO2 values during motion as well with different levels of sensitivities depending on the type of patient monitored	Masimo technology or its equivalent is acceptable
		Scope of supply not included	<u>SCOPE OF SUPPLY:</u> SpO2 probe (per equipment): Neonatal-1 Pediatric/Adult-1 NIBP cuff: Adult- 2 Pediatric- 2 Neonatal (different sizes)- 1 pack each. IBP transducer set - 20 5 lead ECG: 1 per equipment. Temperature probe dual: 5 EtCO2: 3 with standard accessories
		Additional specifications	1. Should have upgradeable module for Cardiac Output measurement using thermo-dilution method of measurement 2. Should be upgradeable to Continuous non-invasive arterial pressure monitoring module (optional) 3. Display of Anaesthesia ventilator data like wave forms for flow, pressure, agent and loops, and trends should be possible with help of MIB cable 4. Should be CE/ US FDA approved. 5. Monitor should be upgradeable to use along with Anesthesia

*R.No. 1232/CC
18/04/17*

[Signature]
CHAIRMAN
 Deptt. of Cardiothoracic Surgery
 J.N. Medical College
 A.M.U., Aligarh

[Signature]
Dr. Mohd. Azam Haseen
 Associate Professor
 Dept. of Cardiothoracic Surgery
 J.N. Medical College
 A.M.U., ALIGARH

[Signature]
Medical Superintendent
 J.N. Medical College Hospital
 A.M.U., Aligarh

