

Pediatric Cardiac ICU equipments

Tender No: JNMC Cardiac ICU(P)/Tender/05/2017-18 dated 25/03/2017

Pre- bid Meeting held at office of Medical Superintendent on 03/04/2017 at 3:00 p.m on this tender

Recommendation:

Establishment of state of the art dedicated Pediatric Cardiac ICU with all the latest equipments and machines.

A. PEDIATRIC CARDIAC ICU

S NO	ITEM	QUANTITY	EMD
1	ICU bed (Fully motorized/five function with separate degree function)	6	Rs10,000
2	ICU bed (Mechanical/five function with separate degree function)	6	Rs 3,000
3	Portable 12 channel ECG recorder	1	Rs 1,200
4	Bedside locker	10	Rs 1,500
5	Neonatal Bubble CPAP	1	Rs 10,000
6	Modular multi para-monitors with Central monitoring system	12	Rs 1,30,000
7	Biphasic Defibrillator	2	Rs 2,000
8	Crash cart	2	Rs 200
9	Emergency/Recovery trolley	1	Rs 160
10	Neonatal/Pediatrics HFO ventilator	1	Rs 36,000
11	Neonatal/Pediatrics ventilator (Turbine based)	1	Rs 24,000
12	Neonatal/Pediatric ventilator (Compressor based)	2	Rs 70,000

Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh



13	Infant radiant warmer	2	Rs 20,000
14	Syringe infusion pump	50	Rs 30,000
15	Volumetric infusion pump	8	Rs 6,000
16	Patient warmer	2	Rs 1,600
17	Fluid warmer	5	Rs 2,500
18	Procedure light	1	Rs 12,000
19	High flow oxygen delivery system	2	Rs 10,000
20	Mechanical chest percussion device	1	Rs 140
21	Suction machine	2	Rs 500

S No	ITEM	EXISTING SPECIFICATIONS	TO BE READ AS
1	ICU BED (FULLY MOTORISED/ FIVE FUNCTION WITH SEPARATE DEGREE FUNCTION)	<p>LAST POINT</p> <p>Product should be ISO 9001, ISO 13485, ISO 14001, OHSAS 18001, CE European, Anti microbial Copper, WHO-GMP and BIFMA</p>	<p>Product should be ISO 9001, ISO 13485, ISO 14001, OHSAS 18001, CE European, Anti microbial Copper.</p>
2	ICU BED (MECHANICAL/ FIVE FUNCTIONS WITH SEPARATE DEGREE FUNCTION)	<p>POINT NO 5</p> <p>The base frame should be mounted on 150 mm dia non-rusting castor twin-wheels with central and directional locking mechanism and pedal operated at the foot end of the bed.</p>	<p>The base frame should be mounted on 150 mm dia non-rusting castor single/twin-wheels with central and directional locking mechanism and pedal operated at the foot end of the bed.</p>
		<p>POINT NO 8</p>	

Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh



