

# Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT  
J.N. MEDICAL COLLEGE HOSPITAL  
ALIGARH MUSLIM UNIVERSITY,  
ALIGARH

No. Ext. 325...../MCH/NAZEER

Dated... 15-04-17

M/s.....

R.No. 1227/CC  
17/04/17

Sealed Quotations/Bids (One Technical and One Financial separately) are invited from manufacturers or their authorized dealers for the following equipment which should be sealed by the tenderers in separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed on the terms and conditions mentioned below on page No. 02 if the outer envelope is not sealed and marked properly as above, the purchase will not assume any responsibility for its misplacement, premature opening, late opening etc.

Quotations should reach this office on or before 08.05.2017 by 03:00 PM.

## Specification for Latest Premium Fully Digital Color Doppler System

- SHOULD BE OF LATEST TECHNOLOGY
- SHOULD HAVE AT LEAST 15" LCD/TFT COLOR DISPLAY
- SHOULD SUPPORT B,M,COLOR,PW,POWER DOPPLER MODES
- Directional power Doppler and advance dynamic flow/e flow MUST BE PRESENT for small flow detection
- SHOULD HAVE 3 ACTIVE PORTS
- SHOULD HAVE HIGH FRAME RATES MORE THAN 700 FPS
- SYSTEM DYNAMIC RANGE SHOULD BE MORE THAN 170DB
- SHOULD HAVE TISSUE HARMONIC IMAGING CAPABILITY IN ALL PROBES
- SHOULD HAVE REAL TIME COMPOUND IMAGING TECHNOLOGY
- BOTH REAL TIME AND FROZEN ZOOM UPTO 16 TIMES
- SHOULD HAVE INTEGRATED TOUCH SCREEN/PLASMA TOUCH SCREEN FOR USER FRIENDLINESS
- REAL TIME QUANTIFICATION OF DOPPLER RARAMETERS WITH SMART TRACE
- FACILITY OF PANAROMIC EIEW/EXTENDED FOV MUST BE PRESENT AS OPTION
- SPECIAL SOFTWARE FOR INTIMA MEDIA THICKNESS CALC (IMT PACKAGE)
- ALL PROBES SHOULD BE MULTIFREQUENCY WITH MINIMUM 5 FREQUENCY SELECTIONS
- CINE LOOP MEMORY OF MORE THAN 15000 FRAMES AND 60 SEC M/D SCROLL
- SCANNING DEPTH MINIMUM 30 CMS
- EXHAUSTIVE SOFTWARE FOR WHOLEBODY APPLICATIONS WITH REPORT FORMATS
- 1000 PATIENT DATA MEMORY SHOULD BE AVAILABLE
- SHOULD HAVE INTEGRATED HARD DISK FOR IMAGE STORAGE/RECALL WITH COMPLETE IMAGE MANAGEMENT
- DIRECT COMPATIBILITY TO ATTACH INKJET/LASERJET PRINTER ALONG ITH A CD-RW MUST BE AVAILABLE
- ANGULAR M-MODE/ Anatomical M-mode function IN Real time with 3 M-mode cursor both in real time and on Frozen Image M-mode cursor rotation of Free & rotate 360 degree, should be in system.
- TISSUE DOPPLER IMAGING SHOULD BE AVAILABLE

## PROBE & ACCESSORIES



1. CONVEX PROBE FREQUENCY RANGE 2TO 6 Mhz- with Tissue Harmonic Imaging. - 01No.
2. LINEAR PROBE should be 5 Mhz to 13 Mhz with maximum wide of 36 mm. - 01 No.
3. ONLINE UPS-2KVA- 01No.
4. Latest Thermal Printer standard make suitable for Equipment. - 01No.

  
Medical Superintendent  
J.N. Medical College Hospital  
A.M.U., Aligarh

Continued on page No. 02

**Terms & Conditions:**

1. Rate & Tax detail **MUST BE** Clearly mentioned in the quotations with printed price list of the manufacturing firm.
2. Only typed quotations/tenders bearing Sales Tax & TIN Nos. on original letter head will be entertained.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt./Public sector institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and Commissioning will be the responsibility of the supplier. After sale service must be provided.
7. The authorised dealer certificate must be attached with quotation.
8. The Equipment/Instruments should be rust proof, made from surgical grade stainless steel.
9. The Equipment/Instruments should be FDA or CE, UL or BIS approved product.
10. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
11. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist and the job description of the hospital technician and company service engineers should be clearly spelt out.
12. Comprehensive on site warranty 03 years + 02 years CMC

  
Medical Superintendent  
J.N. Medical College Hospital  
A.M.U., Aligarh  
  
Medical Superintendent  
J.N. Medical College Hospital  
A.M.U., Aligarh