

Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext 7952 /MCH/NAZEER

R.No - 584/CC
27/03/17

Dated 27/3/2017

M/s.....

Sealed Quotations/Bids (One Technical and One Financial separately) are invited from manufacturers or their authorized dealers for the following equipment which should be sealed by the tenderers in separate covers duly super scribed and both these sealed covers are to be put in a bigger cover, which should also be sealed and duly super scribed on the terms and conditions mentioned below if the outer envelope is not sealed and marked properly as above, the purchase will not assume any responsibility for its misplacement, premature opening, late opening etc.

Quotations should reach this office on or before 17.04.2017 by 03:00 PM.

Specifications for ICU Ventilators

- ICU Ventilators for Neonatal, Pediatric and Adult Patients
- Invasive and Non-invasive modes
- Tidal Volume 2ml to 2000ml
- Modes – CMV, PCV, SIMV-PC, SIMV-PS, SPONTANEOUS, PRVC, CPAP, BIPAP
- Inspitatory hold, Pneumatic Nebulizer, Apnoea backup, Leak compensation, Alarms
- Touch screen display – 7 inch or above with scalars and loops display
- Rechargeable battery
- Compressor based (preferably) or turbine based

Terms & Conditions:

1. Rate & Tax detail **MUST BE** Clearly mentioned in the quotations with printed price list of the manufacturing firm.
2. Only typed quotations/tenders bearing Sales Tax & TIN Nos. on original letter head will be entertained.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt./Public sector institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and Commissioning will be the responsibility of the supplier. After sale service must be provided.
7. The authorised dealer certificate must be attached with quotation.
8. The Equipment/Instruments should be rust proof, made from surgical grade stainless steel.
9. The Equipment/Instruments should be FDA or CE, UL or BIS approved product.
10. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
11. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist and the job description of the hospital technician and company service engineers should be clearly spelt out.
12. Comprehensive on site warranty 05 years.

Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh
Medical Superintendent
J.N. Medical College Hospital
A.M.U., ALIGARH