

OFFICE OF THE MEDICAL SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH-202002 (UP)
Phone No.0571-2720016/2721113/2721114

No. Ext 2067 /MCH

Dated 29.12.2016

To,

R.No. 36/CC
03/01/17

31

Notice Inviting Tender/Quotation

Sealed quotations/tenders are invited for the following work on the terms & conditions printed over leaf. Quotation should reach this office on or before 14.01.2017 by 5:00 P.M

Fabrication of ALS on Force Traveller Ambulance Model WB3350, BSIV

Specifications

- i) The interior paneling of the Patient compartment including sidewalls, partition between Procedure cabin, driver's cabin, roof, door panels and all other surfaces in the patient compartment should be made from long life superior quality Non halogenated FRP of grade C-1196 HF as per NCD unless mentioned otherwise.
- ii) The insulating material should be non-toxic, non-settling type, vermin proof, mild dew proof and non-hygroscopic. The Non halogenated FRP of grade C-1196HF as per NCD wherever used, should have the following Characteristics:
Thickness - minimum 3.0 mm, Inbuilt colour, Fire retardant as per IS - 6746 of 1988 or latest equivalent as applicable.
- iii) The fabrication & fittings of the specified equipments shall permit accessibility for servicing/replacement and adjustment of components/parts and accessories with minimum disturbance to other components and systems.
- iv) The bidder shall ensure that sufficient reinforcement is provided to protect the components, assemblies, pipelines, tubing, wirings, etc which are susceptible to damage / hazards encountered during on-road operations of ambulance. Sufficient reinforcement for holding the wall mounted equipment securely while in transit should be present on the side walls.
- v) Unobstructed access & full functionality of the fittings/equipment as required for optimal patient care must be ensured in this compartment.
- vi) Adequate provision for storage of medicines/consumables/equipment should be made by providing lockable cabinets & drawers. These should be made from fire retardant material in sync with the internal look and feel.
- vii) The drawers should be on guide ways and be provided with appropriate self-restraining mechanism to arrest the inadvertent opening of the unlocked drawers unless pulled while the vehicle is in motion.
- viii) There should be at least 04 numbers of 230 V marked receptacles (each with a switch & a socket with combination of 5/15 AMPS) and one receptacles for 12V DC, of reputed make meeting IS1293 standards.

