

Notice Calling for Quotation/Tender

OFFICE OF THE MEDICAL SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY
ALIGARH.

No. Ext. 1976/MCH

Dated 21/12/16

M/s.....

.....

.....

Sealed Quotations are invited from manufacturers or their authorized dealers for the following item on the terms and conditions mentioned below. Quotations should reach this office on or before 04-01-2017 by 03:00 PM.

Hydraulic Manual OT Table:

- Table should have minimum 3 Sections ;
- It should have following:
 - Mattress for complete Table top in section,
 - A pair of arm bands with pad & fixing clamp.
 - A pair of padded lateral support with clamp.
 - A pair of leg crutches (legging) support.
 - Anaesthetic screen frame with clamp
 - Minimum Length of 765mm.
 - Should have sealed Hydraulic mechanism to avoid oil spillage.
 - Should have atleast 25° Trendelenberg & reverse Trendelenberg.
 - Should have 20° Lateral Tilt movement.
 - Should have minimum 80° back up movement with gas saving mechanism.
 - Head section should have up & down movement.
 - Leg section should have 90° down movement with detachable cover anchor.
 - Should have heavy & sturdy base & compact to provide adequate foot movement to operation team.
 - Table should be mounted on heavy duty casters which offers enhanced weight bearing capacity & free mobility.
 - Should have a minimum patient weight bearing capacity of 100kg.
 - Should have single lever foot operated back pedal.

Terms and Conditions

1. Rate and Tax detail MUST BE clearly mentioned in the quotation
2. Only typed quotations /tenders bearing Sales Tax & Tin No on original letter will be entertained.
3. Cutting /Over writing on quotations /tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should provided, Photocopy of purchase order along with terms and conditions of contract received from any Govt./Public Sector Institution in last 3 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier.
7. After sale service should be provided.
8. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL OR BIS approved product.
9. The equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality
10. User/Technical/Maintenance manuals to be supplied in English.
11. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist.
12. The job description of the hospital technician and company service engineers should be clearly spelt out.
13. Comprehensive on site warranty 03 years

Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh

Medical Superintendent
J.N. Medical College Hospital