

Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext 1949 /MCH/NAZEER

Dated 15-12-2016

M/s.....

Sealed quotations are invited from manufacturers or their authorized dealers for the following **Laprosopic instruments/devices compatible to Laprosopic set of KARL-STORZ, GERMANY** on the terms and conditions as mentioned below:

Quotation should reach this office on or before 31/12/2016 by 03:00 P.M.

1. IMAGE1 H3-LINK, link module, for use with IMAGE 1 FULL HD three- chip camera heads, power supply 100-120 VAC/200 -240 VAC, 50/60 Hz including: Mains Cord, length 300cm Link Cable, length 20cm for use with IMAGE 1 CONNECT TC
2. HOPKINS®II Telescope 30
3. Cold Light Fountain Power LED 175 SCB, with integrated SCB, high- performance LED and one KARL STORZ light outlet, power supply 110-240 VAC, 50/60 Hz consisting of:20161420-1 Cold Light Fountain Power LED 175 SCB 400 A Mains
4. Fiber Optic Light Cable, extremely heat-resistant, safety lock, diameter
5. 26" FULL HD Monitor, wall-mounted with VESA 100 adaption, color systems PAL/NTSC, max-screen resolution 1920x1080, image format 16:9, Video Composite, Video outputs: DVI, 3G-SDI
6. Monitor Stand, suitable for 26" and other monitors, basic monitor stand, tiltable, rotation +/-30, disinfectable, color white, e.g. for use with 26" FULL HD Monitor 9826 NB
7. Trocar, size 11 mm, color code: green, consisting of: 30103 P Trocar only, with pyramidal tip 301303 H2 Cannula without valve, with insufflations stop-cock, length 10.5
8. Trocar, size 6mm, consisting of:30160 P Trocar only, with pyramidal tip 30160 H2 Cannula without valve, with insufflate ion stop-cock, length 10.5
9. Reducer 11/5 mm
10. VERESS pneumoperitoneum Needle with spring-loaded blunt stylet, LUER-lock
11. Clip Applicator, for use with Ethicon Titanium-clips LT-300(medium large), dismantling, rotating, with ratchet to lock the jaw part holding the clip, size 10mm, length 36 cm, consisting of: 30444 H Metal Handle, with ratchet.

Terms & Conditions:

1. Rate & Tax detail MUST BE Clearly mentioned in the quotations
2. Only typed quotations/tenders bearing Sales Tax & TIN Nos. on original letter head will be entertained.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should be provided Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier.
7. After sale service should be provided.
8. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL or BIS approved product.
9. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. User/Technical/Maintenance manuals to be supplied in English.
11. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist.
12. The job description of the hospital technician and company service engineers should be clearly spelt out
13. Comprehensive on site warranty 03 years.


Medical Superintendent
J.N. Medical College Hospital
A.M.U. Aligarh