

Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext 1776...../MCH/NAZEER

Dated 16/11/16.....

M/s.....

Sealed quotations are invited from manufacturers or their authorized dealers for the following item on the terms and conditions mentioned below:


Quotation should reach this office on or before 30/11/2016 by 03:00 P.M.

Cryo cautery machine with Nitrogen cylinder with regulator

- Operating Pressure Range: 40-60 bar
- Coolant: N₂O or CO₂ in two cylinders (A type)
Gas consumption for freezing: ca.35g -50 g/min
Max. exhaust gas volume: 40-60 l/min
- The unit should have Manometer to monitor operating pressure
- A different indicator lamp to indicate freezing and defrosting phase
- Should have a connection pipe for gas exhaust
- It should be mounted in a cart with cylinder case for easy mobilization
- Activation should be via footswitch
- Min. freezing temperature should reach within 5 seconds
- It should be supplied with multiple different sized probe-tips to cater for cervical cryocautery of lesion of all sizes
- All cryoprobes and accessories should be autoclavable
- Should be CE or USFDA approved
- Minimal maintenance, Flawless performance, Available with knob to regulate pressure

Terms & Conditions:

1. Rate & Tax detail MUST BE Clearly mentioned in the quotations
2. Only typed quotations/tenders bearing Sales Tax & TIN Nos. on original letter head will be entertained.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should be provided
Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier.
7. After sale service should be provided.
8. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL or BIS approved product.
9. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. User/Technical/Maintenance manuals to be supplied in English.
11. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist.
12. The job description of the hospital technician and company service engineers should be clearly spelt out.
13. Comprehensive on site warranty 03 years.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh
Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh