

# Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT  
J.N. MEDICAL COLLEGE HOSPITAL  
ALIGARH MUSLIM UNIVERSITY,  
ALIGARH

No. Ext 1593 /MCH/NAZEER

Dated 26/10/2016

M/s.....

Sealed quotations are invited from manufacturers or their authorized dealers for the following Items on the terms and conditions mentioned below:

Quotation should reach this office on or before 15/11/2016 by 03:00 P.M.

S. No. ITEMS

**01. Automatic Component Extractor**

- Should be an open system compatible with any brand of blood bag (Top & bottom as well as Top & Top outlet bags) made to international standards (DB, TBS, QBS).
- Should have more than 10 user selectable optical sensors to separate the components very accurately.
- The equipment comes integrated with Five Radio Frequency sealing heads and clamps.
- The equipment should have built in 3 electronic weighing scales to measure Plasma, SAGM and Buffy coat.
- The equipment has a facility to remove the air from the primary bag before plasma separation.
- The equipment have 11 programs for separation.
- The equipment should be CE marked.
- Should operate with pneumatic controls with noise free compressor.
- Comprehensive on site warranty 5 years.

**02. Blood Collection Monitor**

- Operation – continuous Operation with intermittent loading.
- Battery – 12 V, 2.3 Ah Rechargeable Sealed Maintenance Free Lead Acid Battery.
- Oscillation details -12± 2rpm motor driven oscillation.
- Clamping mode - Motor activated clamping.
- Volume setting range – In steps of 5 ml from 0 ml to 6 ml. automatic storage 8 recall of set volume.
- Display – 16x2 line character Backlit LCD display.
- Time measurement – Time of collection is indicated at the end of every collection.
- Accuracy - ± 2% of the programmed weight.
- Comprehensive on site warranty 3 years.

**03. Tube Sealer**

- Operation – Continuous operation.
- Tube detection – Automatic.
- Maximum diameter of tube that can be sealed – 6mm.
- Sealing time 1-3 sec.
- Comprehensive on site warranty 3 years.

## Terms & Conditions:

1. Rate & Tax detail MUST BE Clearly mentioned in the quotations
2. Only typed quotations/tenders bearing Sales Tax & TIN Nos. on original letter head will be entertained.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should be provided  
Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier.
7. After sale service should be provided.
8. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL or BIS approved product.
9. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. User/Technical/Maintenance manuals to be supplied in English.
11. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist.
12. The job description of the hospital technician and company service engineers should be clearly spelt out.
13. Comprehensive on site warranty 03 years.

  
Medical Superintendent  
J.N. Medical College Hospital  
A.M.U., Aligarh