

Notice Calling for Quotation/Tender

OFFICE OF THE SUPERINTENDENT
J.N. MEDICAL COLLEGE HOSPITAL
ALIGARH MUSLIM UNIVERSITY,
ALIGARH

No. Ext 1512...../MCH/NAZEER

Dated 21-10-2016.....

M/s.....

.....

.....

Sealed quotations are invited from manufacturers or their authorized dealers for the following Items on the terms and conditions mentioned below:

Quotation should reach this office on or before 29/10/2016 by 03:00 P.M.


Fully Fowler's Bed with SS Railing & Section Mattress

Specification:-

- Overall Size (Approximate) 2090mm L x 940mm W x 600 x 600 mm H.
- The main frame will be made from 50mm x 30mm x 16 G ERW rectangular tubes.
- Four sections top will be made from uniformly Oval type perforated 0.9mm C.R.C.A. sheets and will be suitably fitted to the main frame.
- Back-rest and knee rest will be maneuvered by the screw mechanism with Foldable handle.
- Detachable Head and Leg Bows made of Heavy ABS with Inserted Stainless Steel Tube fitting with locking and corner buffer.
- The base frame will be fitted with swivel castor wheels of 125mm dia, 2 with brake, 2 without brakes castors housing and wheels made from high grade non floor staining synthetic materials with integrated thread guards, wheel centre having precision ball bearing to run smoothly.
- The bows will be mounted to the bed frame on welded brackets in such a way that bolt or nut will not appear on top surface of the bed frame.
- Four IV Rod Locations on Bed Frame with one Telescopic IV Rod with 4 Hooks.
- A mattress suitable for the bed made of 10 cm thick 50 density, PU form Covered with rexine closed with Zip at one end.
- The sides will be fitted with Collapsible railing Made of SS horizontal tubes, 18 G fixed with Aluminium alloy end brackets, easily detachable.
- Finishing Steel welding wherever required is be done by either by using MIG welding or Spot welding process to minimize distortion.
- All mild steel components will be pretreated chemically and metal surface will then be coated with epoxy polyester powder with paint film thickness of 60 microns (minimum) and oven baked at 180-200 deg. C.

Terms & Conditions:

1. Rate & Tax detail MUST BE Clearly mentioned in the quotations
2. Only typed quotations/tenders bearing Sales Tax & TIN Nos. on original letter head will be entertained.
3. Cutting/over writing on quotations/tenders will not be accepted.
4. The firm must be registered supplier in any Govt./Semi Govt. or in A.M.U. Aligarh.
5. List of users for the quoted item with contact no. should be provided
Photocopy of purchase order along with terms and conditions of contract received from any Govt/Public sector institution in last 2 years for supply of offered equipment must be enclosed with the price bid.
6. Installation and commissioning will be the responsibility of the supplier.
7. After sale service should be provided.
8. The instruments should be rust proof, made from surgical grade stainless steel and should be FDA or CE, UL or BIS approved product.
9. The Equipment/Instruments might be called for demo, and approval. It is sole discretion of the department to approve or disapprove the quality.
10. User/Technical/Maintenance manuals to be supplied in English.
11. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist.
12. The job description of the hospital technician and company service engineers should be clearly spelt out
13. Comprehensive on site warranty 03 years.


Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh
Medical Superintendent
J.N. Medical College Hospital
A.M.U., Aligarh