

R.No. 867
23-7-13

D.No. 784/BO7
20.7.13.

Department of Botany
Aligarh Muslim University
Aligarh

The Director
Computer Centre
AMU, Aligarh

Through: *Chairman, Department of Botany*

Notice: Corrigendum of tender for purchase of GCMS and RT PCR in Department of Botany, AMU, Aligarh (Ref. 10/MA dt. 13.07.13 and 11/MA dt. 13.07.13)

Ref. No.:14/MA

Dated:.....19.07.13.

This is in reference to our letter no. Ref. 10/MA dt. 13.07.13 and 11/MA dt. 13.07.13 regarding the purchase of "Gas Chromatography Mass Spectrophotometer" and "Real Time PCR"


The following amendments need to be incorporated in the invited tender.

For GCMS:

- Maximum Temperature ramp rate should not be less then 150/120 Degree C/min.
- PTV Injector: Fully Electronic flow/pressure control with pressure programmable range upto 150/100 psi or better. Maximum temperature should be up to 450 Degree with split range of 9000/7500: 1 or better.
- Source temperature should be from 100 Degree to 150 Degree.
- Mass range should be up to 1000 amu or more.
- Turbo molecular pump with minimum capacity of 350/250 L/sec.
- EI Full Scan: 1 pg OFN from m/z 50 to 300 from m/z 272 the S/N should not be less than 600:1.


Terms and conditions: (for both GCMS and RT PCR)

Bids : Two Bids (Separate financial & technical Bid)
FOR : Department of Botany, AMU, Aligarh
Custom duty: University will provide custom duty exemption certificate


(Professor M. Anis)
Coordinator DST FIST- II
D/o Botany, AMU, Aligarh

19.7.13.

Forwarded


20.7.13

Chairman
Department of Botany
A. M. U. ALIGARH
INDIA